

# LA-UR-24-30248

Approved for public release; distribution is unlimited.

**Title:** AI 856 NSR Permit No. 632-M1 Target Fabrication Facility Beryllium  
Machining Lathe TA-35-213-2 Stack Test Report 9-2024

**Author(s):** Carretti, Vincent Anthony

**Intended for:** Environmental Regulatory Document  
**Issued:** 2024-09-26 (rev.1)



Los Alamos National Laboratory, an affirmative action/equal opportunity employer, is operated by Triad National Security, LLC for the National Nuclear Security Administration of U.S. Department of Energy under contract 89233218CNA000001. By approving this article, the publisher recognizes that the U.S. Government retains nonexclusive, royalty-free license to publish or reproduce the published form of this contribution, or to allow others to do so, for U.S. Government purposes. Los Alamos National Laboratory requests that the publisher identify this article as work performed under the auspices of the U.S. Department of Energy. Los Alamos National Laboratory strongly supports academic freedom and a researcher's right to publish; as an institution, however, the Laboratory does not endorse the viewpoint of a publication or guarantee its technical correctness.



## Memorandum

Environmental Protection and Compliance Division  
Compliance Programs Group

**To:** Sarah S. Holcomb, EPC-CP, MS K490  
**Through:** Heather C. Seus, EPC-CP, *HCS* MS J978  
**From:** Vincent A. Carretti, EPC-CP, MS *VAC* J978  
**Phone:** 505-665-1658  
**Symbol:** EPC-DO: 24-274  
**LA-UR:** 24-30248  
**Date:** SEP 25 2024

**Subject: AI 856 NSR Permit No. 632-M1 Target Fabrication Facility Beryllium Machining Lathe TA-35-213-2 Stack Test Report 9-2024**

Los Alamos National Laboratory (LANL) is submitting the final report for the Target Fabrication Facility Beryllium Machining Lathe TA-35-213-2 Stack Test completed August 20-21, 2024. The testing was completed by a subcontractor, Compliance Services and Testing, LLC. The Emission Stack Test is required per NSR Permit No. 632-M1 Condition A600.B "Monitoring." The stack test results show that the emissions are below the permit limit. This report will be submitted through the NMED-AQB Secure Extranet Portal and in the Semi-Annual Monitoring Report for July-December 2024.

Attachment(s): Attachment 1 - AI 856 NSR Permit No. 632-M1 Target Fabrication Facility Beryllium Machining Lathe TA-35-213-2 Stack Test Report 9-2024

Copy: Derek W. Schmidt, Triad, MST-7, [dwschmidt@lanl.gov](mailto:dwschmidt@lanl.gov)  
Christopher T. Wilson, Triad, MST-7, [cwilson@lanl.gov](mailto:cwilson@lanl.gov)  
Terrance J. Akins, Triad, STO-WPGC, [takins@lanl.gov](mailto:takins@lanl.gov)  
Lydia E. Martinez, Triad, EPC-CP, [lydia@lanl.gov](mailto:lydia@lanl.gov)  
Mona M. Michaelis, Triad, EPC-CP, [mmichaelis@lanl.gov](mailto:mmichaelis@lanl.gov)  
Triad, EPC-CP Title V Permit File  
Triad, EPC-CP Correspondence File  
[epccorrespondence@lanl.gov](mailto:epccorrespondence@lanl.gov)  
[eshq-dcrm@lanl.gov](mailto:eshq-dcrm@lanl.gov)

# **ATTACHMENT 1**

AI 856 NSR Permit No. 632-M1 Target  
Fabrication Facility Beryllium Machining  
Lathe TA-35-213-2 Stack Test Report 9-  
2024

EPC-DO: 24-274

LA-UR: 24-30248

Date: SEP 25 2024



New Mexico Environment Department  
 525 Camino de los Marquez, Suite 1  
 Santa Fe, NM 87505  
 Phone (505) 476-4300



Version 11/18/2022

NMED USE ONLY	
DTS	
TEMPO	

## UNIVERSAL STACK TEST NOTIFICATION, PROTOCOL AND REPORT FORM

NMED USE ONLY	
Staff	
Admin	


**Submit to: AQBCR**

I. DATABASE HEADER INFORMATION (drop down menus in bold)			
a. AI# <b>856</b>	Test Report		Initial Compliance Test
d. Company Name: <b>Los Alamos National Laboratory</b>		e. Facility Name: <b>TA-35-213 Target Fabrication Facility</b>	
f. Emission Unit Numbers: <b>TA-35-213-2</b>		g. Emission Unit Description (boiler, Waukesha 7042, etc) <b>Beryllium Machining Lathe</b>	
h. Reports - Tracking Number from notification response: <b>CMT</b>		i. Proposed Test Date: <b>Week of 6-24-24</b>	j. Actual test date: <b>8/20-21/24</b>
k. Reason for test (name permit requirement, NSPS, MACT, consent decree, etc. Indicate here is this notification is a revised test date only) <b>NSR Permit No. 632-M1 Condition A600.B "Monitoring"</b>			

II. GENERAL COMPANY AND FACILITY INFORMATION					
a. Company Address: <b>P.O. Box 1663, MS K490</b>			k. Facility Address: <b>P.O. Box 1663, MS J978</b>		
b. City: <b>Los Alamos</b>	c. State: <b>NM</b>	d. Zip: <b>8 7 5 4 5</b>	l. City: <b>Los Alamos</b>	m. State: <b>NM</b>	n. Zip: <b>87545</b>
e. Environmental Contact: <b>Sarah Holcomb</b>	f. Title: <b>Group Leader</b>		o. Facility Contact: <b>Heather Seus</b>	p. Title: <b>Team Leader</b>	
g. Phone Number: <b>(505) 396-0866</b>	h. Cell Number: <b>(505) 396-0866</b>		q. Phone Number: <b>(505) 412-8832</b>	r. Cell Number: <b>(505) 412-8832</b>	
i. Email Address: <b>sholcomb@lanl.gov</b>			s. Email Address: <b>heatherseus@lanl.gov</b>		
j. Title V Permit Number: <b>P100-R2M5</b>			t. NSR Permit Number: <b>632-M1</b>		
u. Detailed driving directions from nearest New Mexico town: <b>Enter Los Alamos National Laboratory from East Jemez Road. Turn Left on Diamond Drive. Turn left at Pajarito Road. Turn left on Pecos Drive. Building TA-35-213 is approximately 0.5 miles on the left.</b>					

III. TESTING FIRM			
a. Company: <b>Compliance Services &amp; Testing</b>		g. Contact: <b>Chris Spencer</b>	
b. Address 1: <b>7108 Washington St. NE, Ste. A</b>		h. Title: <b>Director</b>	
c. Address 2: <b>PO Box 94191 - 87199</b>		i. Office Phone: <b>505-681-4909</b>	j. Cell Phone: <b>505-681-4909</b>
d. City: <b>Albuquerque</b>	e. State: <b>NM</b>	f. Zip: <b>87109</b>	k. Email Address: <b>cspencer@comptesting.com</b>

# UNIVERSAL STACK TEST NOTIFICATION, PROTOCOL AND REPORT FORM

IV. EMISSION UNIT			STACK PARAMETERS	
a. Emission Unit Number: <b>TA-35-213-2</b>	b. Make & Model Number <b>Moore Nanotech 650FG V2</b>		m. Velocity (ft/sec):	<b>24.3</b>
c. Serial Number: <b>2021-215-UFG</b>	d. Permitted Capacity: <b>0.060 gm/hr Be removal rate</b>		n. Temperature (°C):	<b>70-F</b>
e. Exceptions: Explain if test is late, rescheduled, related to an enforcement action: N/A			o. Stack Diameter, D (in.):	<b>12</b>
			p. Distance to Stack Bends or Obstructions:	
g. Emission Unit Description and brief process name or description: <b>Beryllium Machining Lathe for micro-scale precision machining to create small laser targets for use in the Department of Energy's National Inertial Confinement Fusion (ICF) program.</b>			Upstream, Distance A (in.):	<b>&gt; 1/2D</b>
			Downstream, Distance B (in.):	<b>&gt; 2D</b>
h. Installation Date: <b>March 2024</b>	i. Startup Date: <b>August 20, 2024</b>	k. Date Reached Max. Capacity: <b>August 20, 2024</b>		
I. Control Equipment Description as listed in permit (model, ser. # etc. if applicable): <b>HEPA Filter (HEPA-1)</b>				
			<b>Attach an explanation or drawing to explain any difficult or unusual stack geometry or parameters.</b>	

V. POLLUTANTS AND PROPOSED TEST METHODS		
Pollutant or Parameter:	Proposed Test Methods (Deviations from approved methods require supporting documentation and prior authorization)	Deviation to Test Method Requested
<input type="checkbox"/>	<b>Portable Analyzer Methods for NOx, CO, SO<sub>2</sub></b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>NOx</b> <b>EPA Method 7E</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>CO</b> <b>EPA Method 10</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>SO<sub>2</sub></b> <b>EPA Method 6</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>VOCs</b> <b>(Specify)</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>HAPs</b> <b>(Specify)</b>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<b>PM (TSP)</b> <b>EPA Method 5</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>PM<sub>10</sub></b> <b>EPA Method 201</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>PM<sub>2.5</sub></b> <b>(Specify)</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>Opacity</b> <b>EPA Method 9</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>Visual E.</b> <b>EPA Method 22</b>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<b>Stack Flow</b> <b>EPA Methods 1 - 3</b>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<b>Moisture</b> <b>EPA Method 4</b>	<input type="checkbox"/>
<input checked="" type="checkbox"/>	<b>Other</b> <b>(Specify) EPA Method 29 for Metals - Beryllium</b>	<input type="checkbox"/>
<input type="checkbox"/>	<b>Other</b> <b>(Specify)</b>	<input type="checkbox"/>
List Specific VOC's and HAP's: <b>N/A</b>		

**UNIVERSAL STACK TEST  
NOTIFICATION, PROTOCOL AND  
REPORT FORM**

VI. PROPOSED TEST RUN AND TEST LOAD INFORMATION			
a. Number of Test Runs: <b>3</b>	b. Run Duration <b>240</b>	c. Required by (regulation or permit number): <b>N/A</b>	d. Specific Condition or Section: <b>N/A</b>
PLEASE NOTE – Default run duration is 60 minutes, unless otherwise specified by an applicable regulation.			
e. Expected Load: <b>Maximum available</b>	f. Percent of Permitted Capacity: <b>100%</b>	g. Is this an opacity test? Yes <input type="checkbox"/> No <input checked="" type="checkbox"/>	h. If yes, no. of observation pts.: <b>N/A</b>
i. If expected load during test is less than 90% of capacity, explain: <b>N/A</b>			
<b>NOTE – Failure to test at 90-100% of permitted load will limit unit operation to 110% of tested load until a new initial compliance test is conducted.</b>			
PLANT OR UNIT OPERATING PARAMETERS TO BE MONITORED			
j. List and explain the plant operating parameters that will be monitored and applicable permit conditions or regulatory standards. <b>N/A</b>			

VII. ADDITIONAL DETAILS (where applicable)		
RATA and INSTRUMENTAL ANALYZER CALIBRATION PROCEDURES		
a. Do any of the methods you are proposing utilize instrumental analyzers (i.e.; EPA Methods 3A, 6C, 7E, 10, 18, 25/25A, 320 etc.)? If yes, briefly describe analyzer calibration procedures and/or calibration standard procedures. Enter the highest pollutant concentration expected and the proposed concentrations of calibration gases.	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No
<b>No gaseous sampling of combustion or production byproducts.</b>		
SAMPLING TRAIN LEAK CHECK PROCEDURES		
b. Do any of the methods you are proposing utilize the EPA Method 5 sampling train (i.e.; EPA Methods 1-4, 5, 17, 26/26A, 29, etc.)? If yes, briefly describe sampling train and pitot tube leak check procedures:	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
As described in methods 2 (Pitot tube), 4 (moisture), and 5 (particulate matter).		
EPA METHOD 19 IN LIEU OF EPA METHODS 1-4		
c. Are you proposing to utilize EPA Method 19 in lieu of EPA Methods 1-4?	<input type="checkbox"/> Yes	<input checked="" type="checkbox"/> No
<b>PLEASE NOTE – EPA Method 19 may be utilized in lieu of EPA Methods 1-4, subject to the approval of the Department. If you are proposing to utilize EPA Method 19 in lieu of EPA Methods 1-4, you MUST include a recent fuel gas heating value analysis as well as a recent fuel flow meter calibration certificate, preferably conducted on the day of the test, but no earlier than three months prior to the test date. If the analyses have been conducted prior to the test date, you MUST append the certificates to the protocol. If conducted on the day of the test, you MUST append the certificates to the final test report.</b>		

**UNIVERSAL STACK TEST  
NOTIFICATION, PROTOCOL AND  
REPORT FORM**

<b>VIII. ATTACHMENTS (as needed to support proposed test; check all that apply)</b>	
<b>NOTIFICATION/PROTOCOL ATTACHMENTS</b>	
<input type="checkbox"/>	Road Map Indicating Directions from Nearest New Mexico Town to Facility
<input type="checkbox"/>	Schematic of process being tested showing emission points, sampling sites and stack cross-section
<input type="checkbox"/>	Copy of proposed test methods (except for those promulgated test methods found in 40 CFR 51, 60, 61 and 63)
<input type="checkbox"/>	Fuel Heating Value Analysis
<input type="checkbox"/>	Fuel Flow Meter Calibration Certificate
<input type="checkbox"/>	Other:
<input type="checkbox"/>	Other:
<b>TEST REPORT ATTACHMENTS</b>	
<input checked="" type="checkbox"/>	<b>Section 2. Tables of Results</b>
<input checked="" type="checkbox"/>	<b>Supporting Documents (Specify) Test Report</b>
<b>Retain Report Section 3 - Test Procedures, Data, Calculations, Appendices – 2 years NSR permits, 5 years TV</b>	



**NEW MEXICO ENVIRONMENT DEPARTMENT  
AIR QUALITY BUREAU  
NSR PERMIT NUMBER 632-M1**

**COMPLIANCE TEST REPORT  
ON:  
EXHAUST EMISSIONS**

FROM A:  
**LATHE MACHINE (UNIT #1 AND 2)**

LOCATED AT THE:  
**TA-35-213 BUILDING**

PREPARED FOR:  
**LOS ALAMOS NATIONAL LABORATORY**

TEST DATE:  
**AUGUST 20-21, 2024**

PREPARED BY:  
**COMPLIANCE SERVICES AND TESTING, LLC**

PROJECT NUMBER:  
**2634**



P.O. Box 94191-87199  
7108 Washington St. NE  
Suite A  
Albuquerque, NM 87109  
(505) 681-4909 Phone  
[www.comptesting.com](http://www.comptesting.com)

# Table of Contents

Introduction	1
Table 1:    Background Data	2
Summary of Results	3
Table 2:    Summarized Emission Results	3
Process Description and Sampling Locations	4
Analytical Technique	5-6
Figure 1:    Manual Sampling System Diagram	7
Quality Assurance Activities	8
Appendices	
A:    Tabular Results and Stack Diagram	
B:    Sampling Data Sheets	
C:    Example Calculations	
D:    QA/QC Activities and Equipment Calibration Results	
E:    Laboratory Analysis Report and Chain of Custody	

## **Introduction**

Los Alamos National Laboratory (LANL) contracted Compliance Services and Testing (CST) to perform an emissions test on the existing lathe (#1) and a new lathe (#2) in service at Building TA-35-213. The building is located within the boundary of LANL, located in Los Alamos, New Mexico in Los Alamos County.

The testing was to show compliance with the NSR permit issued to this facility and followed procedures with the NMED protocol. The testing procedures are found in the Code of Federal Regulations, Title 40, Part 60, Appendix A, Method 1 for sampling locations, Method 2 for velocity determination, Method 4 for moisture content (H<sub>2</sub>O), Method 5 for Isokinetic Sampling, and Method 29 for Metals Emissions, in particular, beryllium emissions. The diluent concentrations for oxygen (O<sub>2</sub>) and carbon dioxide (CO<sub>2</sub>) were considered ambient in lieu of Method 3A due to the source being a non-combustion type.

The information contained in this report has been reviewed and approved as being truthful, accurate, and complete.

## Table 1 - Background Data

<b>Source Operator:</b>	<b>Los Alamos National Laboratory (LANL)</b> Attn: Vincent Carretti P.O. Box 1663, MS J978 Los Alamos, NM 87545
<b>Test Contractor:</b>	<b>Compliance Services and Testing (CST)</b> Attn: Chris Spencer 7108 Washington NE Ste. A Albuquerque, New Mexico 87109 Phone: (505) 681-4909
<b>Test Participants:</b>	<b>CST</b> Chris Spencer – Director Matt Magee – Environmental Scientist  <b>LANL</b> Derek Schmidt – Engineering Team Leader
<b>Test Dates:</b>	August 20-21, 2024
<b>Location:</b>	TA-35 is located within Los Alamos National Laboratory, Los Alamos, New Mexico
<b>Test Methods:</b>	<b>Title 40 of the Code of Federal Regulations, Part 60, Appendix A</b> Methods 1,2,4,5 for Isokinetic Sampling Methods Method 29 for Metals Emissions
<b>Regulatory Permit:</b>	NMED-AQB NSR Permit #632-M1

## Summary of Results

The beryllium particulate emissions emitted from two lathes were tested to demonstrate initial compliance with the permit in place for this facility. Even though the calculated beryllium removal rates indicate each lathe was operating at 50% of the permit limit (0.030 gm/hr actual vs 0.060 gm/hr permitted), the lathe operators indicated that the lathes were processing the maximum amount of beryllium achievable, and at a maximum rate achievable, during the testing. As such, it has been determined that the lathes were operating at 100% load/maximum production. The lathes could not operate at a higher amount and rate without risking the beryllium breaking. Additionally, both lathes operated during the entire 4-hour test without downtime. The results of the testing show the particulate matter emissions are below the permit limit. A summary of the test results is listed below.

**Table 2 – Summarized Emission Results**

	<i>Beryllium (Be)</i>	<i>% Load</i>
<b>Lathe Unit #</b>	<b>gm/hr</b>	<b>%</b>
<i>Limits</i>	<b>6.25E-5</b>	<b>≥90%</b>
<b>1 &amp; 2</b>	<b>^2.55E-5</b>	<b>100.0</b>

\*gm/hr = grams per hour

^Value based on minimum detection limit.

Example calculations are included in Appendix C to give examples of how raw data was used to determine stack conditions.

## Process Description and Sampling Location

The emissions pass through a 99.95% efficient HEPA associated with the fume hood used to capture the particulate matter and vent the emissions through the exhaust stack to the atmosphere. The exhaust temperature is at room temperature so neither the probe nor filter were heated. An ice-bathed cooled impinger train was used to collect the gaseous sample content that passed through the filter.

This diagram represents a general process description and is not to scale. This representation is the equipment being utilized and the placement order of that equipment. The values presented are not of stack conditions during the testing event.



The sampling took place from two exhaust stack ports, separated by 90° and placed 18 feet and 21 feet downstream and upstream from the nearest flow disturbances in the exhaust stack. A Method 1 Sampling Traverse Point worksheet is included in Appendix A along with the results of the cyclonic flow test recorded during the preliminary stack flow test.

## Analytical Technique

An emissions test was performed from an exhaust stack associated with the vent of two lathes in use at Los Alamos National Laboratory. The test procedures utilized are outlined in the Code of Federal Regulations, Title 40, Part 60, Appendix A, Methods 1-5 and 29. The sampling time was 240 minutes for each of the three test runs.

**EPA Method 1--Sample and Velocity Traverses for Stationary Sources:** The sampling locations and traverse points met Method 1 distance requirements to enhance the quality of the data obtained. A cyclonic flow check was performed during the preliminary investigation by measuring the yaw angle at each sampling point. The results of that test found no cyclonic flow in the stack, therefore there is no stratification present.

**EPA Method 2--Determination of Stack Gas Velocity and Volumetric Flow Rate (Standard Pitot Tube):** The velocity of the stack gas was constantly measured using a S-type Pitot tube at the sampling points. The Pitot tube is a component of the Method 5 probe assembly probe and is located on the sampling pane along with the thermometer and sampling nozzle. The Pitot measures the velocity of the stack gas by measuring the impact and wake-side pressures and creating a differential reading measured in inches water. The values obtained were factors in calculating the exhaust flow rate and sampling rates.

**EPA Method 4--Determination of Moisture Content in Stack Gases:** The moisture content measurements were made per EPA Method 4 (gravimetric) by pulling stack gas through an impinger train emerged in an ice bath. The impinger train acts as a moisture knockout. The sample then went through a separate stainless steel/Teflon® diaphragm pump and calibrated dry gas meter. The weights of the impingers (four total) were measured using a calibrated scale accurate to 0.1 grams. One moisture sample was collected over a 240-minute period from the stack concurrent with each test run. 100 mL of acidic hydrogen peroxide (5% HNO<sub>3</sub>/10% H<sub>2</sub>O<sub>2</sub>) was placed in the first two impingers and acted as the primary collection media per Method 29.


**EPA Method 5--Determination of Particulate Matter Emissions from Stationary Sources:** The sampling techniques outlined in Method 5 Isokinetic Sampling were used to collect a gaseous sample at the same flow rate as the exhaust flow rate. To avoid over- or under-sampling of particulate matter, the gas needs to be metered in a manner that captures particulates in a representative

manner. The nozzle diameter, the diluent concentrations, stack and meter box temperatures, along with the sample pump flow rate are all factors in determining the isokinetic rate of sampling. The diluent concentration was ambient air.

**EPA Method 29--Determination of Metals Emissions from Stationary Sources:** The sampling techniques outlined in Method 5 were used to collect particulate matter on a filter and gaseous emissions are collected in an aqueous acidic solution of hydrogen peroxide. The recovered samples are digested and analyzed for the target compounds by atomic absorption spectroscopy (AA) or inductively coupled argon plasma optical emission spectrometry (ICAP). The aqueous acidic solution is gravimetrically analyzed to determine the moisture content of the stack before being segregated for analysis. The filters and front-half fractions are collected, desiccated, and weighed before and after the tests in the same manner if TSP emissions are to be determined.



# Figure 1. Method 5 Console Sampling System



## Quality Assurance Activities

The console sampling system was leak checked before and after each test run by demonstrating that a vacuum (10-15" Hg) greater than sampling conditions (0-5" Hg) could be pulled and held for a minute. A leak test was then conducted on each leg of the Pitot tube before use and after use at a pressure greater than 3" for 15 seconds. These tests were conducted to ensure that ambient air had not diluted the sample. If any leaks were found during the pre-test inspection, they were resolved and re-tested until acceptable checks were obtained (<0.02 cf/min). No leaks greater than the allowed were detected on the sampling system during the post-test checks. The dry gas meter incorporated into the metering console was calibrated before the test and verified after per method requirements to ensure that an accurate sample volume was recorded. The console temperature gauges were checked for acceptable tolerance (<2%) prior to and after field use.

The sampling nozzle was measured at three traverse lines before testing to verify nozzle diameter. A stack cyclonic flow check was performed before testing to ensure the gas flow was laminar. The results from that test showed the average of the absolute yaw angle at each sampling point was under 20%, therefore, the stack did not exhibit cyclonic flow. The scale used to weigh the impinges was checked in triplicate with a 1 kg weight and showed variance less than tolerance (<1%). Each filter used was checked for tears or holes by holding each up to a light source to verify integrity. None were found that would comprise the sample collection.

All calibration data and verification checks are contained in Appendix D.

## **Appendix A: Tabular Results and Stack Diagram**

# Summary of Results

## LANL, TA35-213 Lathe #1 and Lathe #2

**Company:** Los Alamos National Laboratory

**Location:** TA-35-213

**Technician:** CS, MM

**Source:** Lathe #1 and Lathe #2

**Model:** 350 UPL (Lathe #1) and 650FGV2 (Lathe #2)

**Source:** Ambient

<i>Test Run Number</i>	<i>1</i>	<i>2</i>	<i>3</i>	
Date	8/20/24	8/20/24	8/21/24	
Sampling Time (min)	240.0	240.0	240.0	
<b><i>Fume Hood Operating Parameters</i></b>				<b><i>Average</i></b>
Load (%)	100.0	100.0	100.0	<b>100.0</b>
<b><i>Constants</i></b>				
Constant K <sub>1A</sub> (L / g)	1.336	1.336	1.336	<b>1.336</b>
Constant K <sub>3</sub> (°R-L / "Hg-cf)	499.7	499.7	499.7	<b>499.7</b>
Conversion Factor ("Hg / "H <sub>2</sub> O)	0.07355	0.07355	0.07355	<b>0.07355</b>
Pitot Tube Constant K <sub>p</sub> (√(lb/lb.mol-"Hg/°R-"H <sub>2</sub> O)	5129.4	5129.4	5129.4	<b>5129.4</b>
STP / Time Constant K <sub>y</sub> (°R-min / "Hg-hr)	1058.8	1058.8	1058.8	<b>1058.8</b>
<b><i>Ambient Conditions</i></b>				
Pressure Altitude (MSL)	7260	7260	7260	<b>7260</b>
Atmospheric Pressure ("Hg)	22.88	22.88	22.88	<b>22.88</b>
Average Dry Bulb Temperature (°F)	75	95	90	<b>86.7</b>
Average Wet Bulb Temperature (°F)	70	83	80	<b>77.7</b>
Humidity (lb/lb air)	0.0190	0.0282	0.0258	<b>0.0244</b>
<b><i>Stack Parameters</i></b>				
Static Stack Pressure (P <sub>g</sub> - "H <sub>2</sub> O)	-0.16	-0.16	-0.16	<b>-0.16</b>
Absolute Stack Pressure (P <sub>s</sub> - "Hg)	22.87	22.87	22.87	<b>22.87</b>
Average Stack Temperature (T <sub>s</sub> - °R)	534.9	535.5	536.9	<b>535.8</b>
Stack Moisture (B <sub>ws</sub> - %)	1.1866	1.2655	1.3209	<b>1.2576</b>
Dry Gas Fraction (1-B <sub>ws</sub> )	0.9881	0.9873	0.9868	<b>0.9874</b>
Dry Stack Gas Molecular Wt. (M <sub>D</sub> - lbs/lb-mole)	28.842	28.842	28.842	<b>28.842</b>
Wet Stack Gas Molecular Wt. (M <sub>s</sub> - lbs/lb-mole)	28.714	28.705	28.699	<b>28.706</b>
<b><i>Stack Velocity and Flow Rate via Pitot Tube</i></b>				
Stack Velocity (V <sub>s</sub> - ft/sec)	32.04	32.90	33.22	<b>32.72</b>
Stack Velocity (V <sub>s</sub> - ft/min)	1922.32	1973.94	1993.04	<b>1963.10</b>
Stack Flow, wet (Q <sub>a</sub> - acf/sec)	25.16	25.84	26.09	<b>25.70</b>
Stack Flow, wet (Q <sub>a</sub> - acf/min)	1,509.78	1,550.32	1,565.33	<b>1541.81</b>
Stack Flow, dry (Q <sub>s</sub> - scf/min)	1,125.62	1,153.48	1,161.10	<b>1146.73</b>
Stack Flow, dry (Q <sub>s</sub> - scf/hr)	67,537	69,209	69,666	<b>68,804</b>
<b><i>Dry Gas Meter Sampling Data</i></b>				
Corrected Metered Volume (V <sub>mcorrected</sub> - CF)	176.700	182.135	189.922	<b>182.919</b>
Volume of DGM Sample (V <sub>mstd</sub> - L)	3681.938	3803.847	3921.707	<b>3802.498</b>
Corrected Metered Volume (DSCF)	130.593	134.872	139.075	<b>134.846</b>
Nozzle Diameter (ft <sup>2</sup> )	3.88E-04	3.88E-04	3.88E-04	<b>3.88E-04</b>
Isokinetic Rate (%) {90 ≤ I ≤ 110}	97.95	98.71	101.12	<b>99.26</b>
<b><i>Beryllium Data</i></b>				
^Sample Mass Collected - Front Half (μg)	0.20	0.20	0.20	<b>0.200</b>
Sample Mass Collected - Back Half (μg)	0.00	0.00	0.00	<b>0.000</b>
*Total Mass Collected (μg)	0.20	0.20	0.20	<b>0.200</b>
Be Concentration (g/DSCF)	1.53E-09	1.48E-09	1.44E-09	<b>1.48E-09</b>
<b>Be (g/hr) {Permit Limit = 6.25E-5}</b>	<b>2.59E-05</b>	<b>2.57E-05</b>	<b>2.50E-05</b>	<b>2.55E-05</b>

g=grams; μg=micrograms

^Sample concentration used is the method detection limit.

\*Back-half fraction is non-detect, equal to zero. Front-half fraction less than blank, therefore equal to zero.

# Sampling Traverse Point Layout

## TA-35-213 - Lathe #1 and Lathe #2

Date: 8/19/24  
 Company: LANL  
 Location: TA-35-213  
 Source: Lathe #1 and Lathe #2  
 Technician(s): CS, MM

Port + Stack ID (in): 16.00  
 Port Extension (in): 4.00  
 Stack ID (in): 12.00  
 Stack Area (ft<sup>2</sup>): 0.785  
 Duct Diameters **upstream** from flow disturbance (A): 21.00  
 Duct Diameters **downstream** from flow disturbance (B): 18.00  
 Total Required Traverse Points: 8  
 No. of Traverse Points per Diameter: 4

**NOTES:**

Stack Diagram						Stack Cross Section		
Unit Information								
Traverse Point Number	Number of Traverse Points on a Diameter					Method 7E Port Length Offset	Method 1 Port Length Offset	Method 5 Port Length Offset
	3	4	6	8	12			
	<i>Percent distances (%)</i>							
1	16.7	6.7	4.4	3.2	2.1			4.80
2	50.0	25.0	14.6	10.5	6.7			7.00
3	83.3	75.0	29.6	19.4	11.8			13.00
4	93.3	93.3	70.8	32.3	17.7			15.20
5			85.4	67.7	25.0			
6			95.6	80.6	33.6			
7				89.5	64.4			
8				96.8	75.0			
9				82.3				
10				88.2				
11				93.3				
12				97.9				
<p>*Stack diameters ≤ 24 in shall have no traverse points located w/in 0.5 in of the stack wall                      **Stack diameters &gt; 24 in shall have no traverse points located w/in 1 in of the stack wall</p>								

# Traverse Point Layout Calculations


## TA-35-213 - Lathe #1 and Lathe #2

**Method 1/7E/PS-2 - Traverse Point Layout Calculations**  
 Enter the required information in the appropriate row/columns and the spreadsheet will calculate the required number of traverse points and their locations on the probe/pitot tube.

<b>Use Method 1 or Method 20 For the Non-Particulate Traverse?</b>		<b>Stack Layout</b>	Stack + Port (in)	Port (in)	Stack ID (in)	Stack Radius (ft)	Stack Area (sq ft)	Distance (ft) Upstream (A)	Distance (ft) Downstream (B)	Upstream (A)	Downstream (B)	Upstream (A)	Downstream (B)	Meets Method 1 Min. Distance Criteria?	<b>Method 1 Calculated</b> Min. No. of Traverse Points	Total # of DIAMETERS Traversed	Traverse Points per DIAMETER
<b>Method</b>		<b>CIRCULAR</b>															
								<i>see Flow Disturbance Diagram below</i>									
<b>1</b>		Non-Particulate															
		Particulate	16	4	12	0.50	0.785	21.00	18.00	21.00	18.0	YES	8	2	4		
		<b>RECTANGULAR</b>	Port (in)	**Length (in)	Width (in)	Eq. Diameter (in)											
		Non-Particulate															
		Particulate															

NOTE: If needed, you can override the calculation of # of traverse points for rectangular stacks in order to achieve an appropriate matrix.

<b>Method 1 Circular Traverse Point Layout</b>						
Traverse Point Number	Non-Particulate			Particulate		
	Traverse Pt. Location (in)	Traverse Pt. Location + Port (in)	*Distance from Stack Wall	Traverse Pt. Location (in)	Traverse Pt. Location + Port (in)	*Distance from Stack Wall
1				0.80	4.80	
2				3.00	7.00	
3				9.00	13.00	
4				11.20	15.20	
5						
8						
9						
10						
11						
12						
13						
14						
15						
16						
17						
18						
19						
20						



\*Correct traverse point distances that are too close to stack wall. Select a minimum of 1.0" or 0.5" for stack IDs of >24" and ≤24", respectively.

Note: If traversing large diameter, 4-port stacks with a short probe, simply mark the probe with the first half of the traverse points and traverse each of the 4 ports (which equals the total # of traverse pts.).

**Appendix B:  
Sampling Data Sheets**

## Pre-Lim Sampling Time, K Factor, Isokinetic Rates TA-35-213 - Lathe #1 and Lathe #2

<b>Stack Moisture &amp; Molecular Wt.</b>		<b>Predicted</b>	
CO2 (%)		0	
O2 (%)		20.90	
N2 (%)		79.10	
DGM Sample Volume (ft3)		180.000	
Ending Impinger Wt. (g)		25.0	
Dry Gas Meter Factor (Kd)		0.993	
Average Dry Gas Meter Temperature (oF)		85	
Atmospheric Pressure (in Hg, abs.)		22.88	
Stack Gas Moisture (% volume)		0.8828	
Dry Gas Fraction		0.991	
Stack Gas Molecular Wt. (lbs/lb-mole)		28.74	
<b>Stack Parameters</b>		<b>Temp.</b>	<b>Yaw Angle</b>
ΔP #1	0.27	73	10.00
ΔP #2	0.31	73	8.00
ΔP #3 (Static Pressure = -0.17)	0.28	74	-5.00
ΔP #4	0.27	74	-12.00
ΔP #5	0.18	76	-4.00
ΔP #6	0.26	74	-1.00
ΔP #7 (Static Pressure = -0.15)	0.31	74	-1.00
ΔP #8	0.25	74	-11.00
<b>Average</b>	<b>0.266</b>	<b>74.0</b>	<b>-2.00</b>
Pitot Tube Factor	0.84		
Sum of Square Root of Vertical Component	4.116		
Number of Traverse Points	8		
Average Square Root of ΔP's	0.515		
Average Temperature (°F)	74.0		
Static Pressure (in. H2O)	-0.16		
Stack Diameter (in.)	12		
Stack Area (ft2)	0.785		
<b>Stack Velocity</b>			
Stack Velocity (ft/sec @ stack conditions)	33.300		
<b>Stack Velocity (ft/min @ stack conditions)</b>	<b>1,998.02</b>		
Stack Velocity (Std ft/min)	1,496.62		
<b>Stack Flow</b>			
Stack Flow, wet (CF/Min)	1,569.24		
Stack Flow, wet (Std CF/Hr)	71,154.95		
Stack Flow, dry (Std CF/Min)	1,175.45		
<b>Stack Flow, dry (Std CF/Hr)</b>	<b>70,526.83</b>		
<b>Isokinetic Nozzle Selection &amp; K Factor Calculation</b>			
Stack Velocity (ft/min @ stack conditions)	1998.02		
<b>Target Sampling Rate (CFM)</b>	<b>0.70</b>		
<b>Target Nozzle Area (Ft2)</b>	<b>3.50E-04</b>		
Selected Nozzle Number (xx/32")	0.27		
Actual Measured Nozzle Area (Ft2)	3.88E-04		
No. of M-5 Traverse Pts	8		
Minutes per point (min.)	30.00		
Sampling Time (min.) (>60 min)	240.00		
Sampling Time (hr)	4.00		
Expected Total Sample Vol. (CF)	185.984		
Expected Sample Vol. (SCF/Hr) (>30 SCF)	33.105		
Sampling Rate (CFM) (using actual nozzle)	0.775		
ΔH ("H2O)(see Meter Box Calib Curve)	<b>1.624</b>		
<b>K-Factor (ΔH/ΔP)</b>	<b>6.13</b>		

Permit Limit (g/hr)	6.25E-05	<b>Nozzle Measurements:</b>	
Sample Rate (ft3/min)	0.775	0.26	inches
Minimum catch size (mg)	0.0001	0.27	inches
Stack Flow Rate (SCF/hr)	70,527	0.27	inches
<b>Minimum Sample Time (min)</b>	<b>145.62</b>	<b>0.000388</b>	<b>ft^2</b>
<b>Minimum Sample Time (hr)</b>	<b>2.43</b>		



## In-Stack Detection Limit and Test Time Determination LANL - TA-35-213 Lathe #1 and Lathe #2

$$\text{TIME} = \frac{\text{Stack Flow} \times \text{Catch Amnt}}{\text{Sample Rate} \times \text{PL} \times \text{Conv Fact}} \quad \frac{\text{scf} \cdot \text{mg} \cdot \text{min} \cdot \text{hr} \cdot \text{g}}{\text{hr} \cdot \text{scf} \cdot \text{g} \cdot \text{mg}}$$

<b>Permit Limit</b>	6.25E-05	g/hr	<b>Catch (mg)</b>	<b>Min</b>	<b>Hr</b>	
<b>Permit Limit</b>	1.38E-07	lb/hr	<b>0.0001</b>	<b>157</b>	<b>2.6</b>	<b>Instrument MDL</b>
<b>ISCL</b>	9.13E-04	ng/scf				
<b>Sample Rate</b>	0.70	cf/min	<b>0.001</b>	<b>1,565</b>	<b>26.8</b>	<b>Reportable Limit</b>
<b>ISCL</b>	0.91271	mg				
<b>Stack Flow Rate</b>	68,483	scf/hr	<b>0.8896</b>	<b>1,392,514</b>	<b>23,208.6</b>	<b>ISCL (mg/scf)</b>
<b>Stack Flow Rate</b>	1,141	scf/min				<b>= PL ÷ SFR x 2.67x10<sup>5</sup></b>
<b>Conv Fact</b>	1000	mg/g				

### In-Stack Detection Limits & Sample Times (NJDEQ)

- 1) If the in-stack detection limit is higher than the level of the standard, compliance cannot be demonstrated.
- 2) Actual in-stack method detection limits (ISDL) are based on actual source sampling parameters and analytical results. Actual detection limits can be improved through increased stack gas sampled (sample time), reducing the total volume of the digested samples, improving the analytical detection limits or any combination of the three.
- 3) The goal of the sample program will be to have the in-stack detection limit at 1/10 the Permit Allowable expressed as the in-stack concentration limit (ISCL). The ratio of ISCL to the ISDL should be >10. The following will detail these determinations and will establish the sampling time.

\*\*\*Based on the preceding, each sample run will be **180** minutes.

### In-Stack Concentration Limit (ISCL): ISCL (ug/scm) = E / F x 2.67E8

Where:

$$E = \text{Permit allowable (lb/hr)} = 1.38E-07 = 6.25E-5 \text{ (g/hr)} \div 453.6 \text{ (g/lb)}$$

$$F = \text{stack flow rate (dscfm)} = 1,141$$

### In-stack Detection Limit: ISDL (ug/scm) = [A (mg/mL) x B (mL) ÷ C (scf)]

Where:

A = analytical detection limit:	A=	0.0001	mg/mL
B = volume of analyte collected:	B=	400	mL
C = volume of stack gas sampled:	C=	41.664	scf

### PARAMETERS

22.88 ATM at MSL

75 Stack Temp

1.25E-04 Permit limit (20% overage)

1.04E-04 estimated g/hr per Lathe

400 Approx. mL of sample

28.3168 L/CF

0.0001 mg/mL ICAP MDL

124.991 SCF / 3 Hr

41.66 SCF / Hr of Sample

1.471 Sample Vol (scm/hr)

ISCL (ug/scm) =	9.13E-04	1 x ISCL > ISDL	2 x ISCL > ISDL	5 x ISCL > ISDL	10 x ISCL > ISDL	Sample Hours
ISDL (ug/scm) =	9.60E-04	FALSE	FALSE	FALSE	FALSE	1
ISDL (ug/scm) =	4.80E-04	TRUE	FALSE	FALSE	FALSE	2
ISDL (ug/scm) =	3.20E-04	TRUE	TRUE	FALSE	FALSE	3
ISDL (ug/scm) =	2.40E-04	TRUE	TRUE	FALSE	FALSE	4
ISDL (ug/scm) =	1.92E-04	TRUE	TRUE	FALSE	FALSE	5
ISDL (ug/scm) =	1.60E-04	TRUE	TRUE	TRUE	FALSE	6
ISDL (ug/scm) =	1.37E-04	TRUE	TRUE	TRUE	FALSE	7
ISDL (ug/scm) =	1.20E-04	TRUE	TRUE	TRUE	FALSE	8
ISDL (ug/scm) =	1.07E-04	TRUE	TRUE	TRUE	FALSE	9
ISDL (ug/scm) =	9.60E-05	TRUE	TRUE	TRUE	FALSE	10
ISDL (ug/scm) =	8.73E-05	TRUE	TRUE	TRUE	TRUE	11
ISDL (ug/scm) =	8.00E-05	TRUE	TRUE	TRUE	TRUE	12
ISDL (ug/scm) =	7.39E-05	TRUE	TRUE	TRUE	TRUE	13
ISDL (ug/scm) =	6.86E-05	TRUE	TRUE	TRUE	TRUE	14
ISDL (ug/scm) =	6.40E-05	TRUE	TRUE	TRUE	TRUE	15
ISDL (ug/scm) =	6.00E-05	TRUE	TRUE	TRUE	TRUE	16

# Isokinetic Worksheet

## TA 35-213 - Lathe #1 and Lathe #2, Run 1

Date: 8/20/24  
 Client: LANL  
 Plant: TA-35-213  
 Emission Source: Lathe #1 and Lathe #2  
 Technicians: CS, MM  
 DGM No: Console 2  
 DGM Factor, Y: 0.993  
 AH@: 1.624  
 Stack Diameter ("): 12  
 Stack Area (ft<sup>2</sup>): 0.785

Critical Isokinetic Test Data			Pre-Test Calibration Check		
Atmosph. Press. ("Hg)	22.88	Static P	V1	635.094	cf
Pilot Tube Factor, Cp	0.84	B	Vf	643.901	cf
Static Pressure (in H2O), P <sub>g</sub>	-0.16	-0.17	Vt	8.807	cf
M <sub>s</sub> - Molecular Weight (lb/lb-mole)	28.71	A	ΔH	1.60	"H2O
Dry Gas Fraction, 1-Bws	0.9881	-0.15	Time	10	min
Nozzle Diameter (in)	0.267	Nozzle (in.)	Im	540	R
Nozzle Area (ft <sup>2</sup> )	3.88E-04	0.26	Phar	22.88	"Hg
Total Sample Time (min)	240	0.27	constant	0.0319	"Hg-clm/R
Number of Points	8	0.27	Yc	0.9852	
Filter #	QF-15	0.267	ΔH@	1.5750	

Dry Gas Meter Leak Checks			Pilot Tube Leak Check		
pre-test	0.000 cf @	15"	+/-	4.8/4.7	
post-test	0.002 cf @	10"	+/-	---	
Moisture Data					
Run 1	Imp 1	Imp 2	Imp 3	Imp 4	Total Wt
Beginning	713.7	689.2	596.9	971.6	2971.4
Ending	714.8	693.0	599.7	997.0	3004.5
Total	1.1	3.8	2.8	25.4	33.1
Gaseous Data (%)					
O2	20.90	CO2	0.04	N2	79.1
Scale Prelim Check (1000g): 1000.0 999.9 1000.0 999.97					

Start Vol	644.322	Stop Vol	822.268	Actual Volume	177.946	Target	180.000	ACF x Y	176.700	SCF	130.593
Start Time	8:26	Stop Time	12:27	Sample Time per Point	30	K Factor	4.5	Z <sup>2</sup> ΔP	3.955	Avg ΔP	0.249

Isokinetic Rate (%)	97.95
---------------------	-------

Sample Point per Port	Start Meter Time	Observed Clock Time	Observed Meter Reading (ft <sup>3</sup> )	Next Point Target (ft <sup>3</sup> )	ΔP ("H2O)	ΔH ("H2O)	Desired ΔH ("H2O)	Observed K Factor	Sample Vacuum ("Hg)	DGM Temp. (°F)	Last Imp. (°F)	Filter Temp. (°F)	Probe Temp. (°F)	Stack Temp. (°F)	Point Vol Sampled (DSCF)	Σ - ΔΔP ("H2O)	Stack Velocity (ft/s)	Average ISO (%)
1	0:00	8:26	644.322	661.627	0.150	0.913	0.68	6.04	0.50	72.0	43.7	72.7	---	73.7	15.101	0.387	25.07	115.4
2	30:00		664.300	687.085	0.250	1.533	1.13	5.96	1.50	83.3	39.7	77.0	---	73.0	18.932	0.500	32.35	113.5
3	60:00		689.830	715.413	0.310	1.900	1.40	5.89	2.00	89.0	42.0	78.7	---	74.0	21.272	0.557	36.06	113.4
4	90:00		718.780	744.218	0.303	1.600	1.37	5.05	1.67	91.3	43.0	78.7	---	74.0	19.627	0.551	35.67	111.2
5	120:00	10:27	745.631	762.941	0.140	0.700	0.63	4.78	0.50	92.0	42.0	79.3	---	76.7	12.761	0.374	24.29	109.7
6	150:00		763.160	786.022	0.243	0.887	1.10	3.48	1.00	93.0	40.7	80.0	---	76.0	14.564	0.493	32.00	105.9
7	180:00		783.190	808.776	0.303	0.900	1.37	2.83	0.63	94.3	42.7	80.0	---	76.0	14.326	0.551	35.73	101.3
8	210:00		802.940	828.128	0.293	0.850	1.32	2.76	0.63	94.7	44.3	80.7	---	75.7	14.010	0.542	35.13	97.9
Finish	240:00	12:27																
Total			177.946	Average	0.249	1.160	1.121	4.599	1.054	88.71	42.25	78.38	---	74.88	130.5927	3.9546	32.038	108.54

BG Nozzle Number	1	2	3	4	5	6	7
Nozzle Diameter (in.)	0.125	0.187	0.250	0.312	0.375	0.437	0.500
Nozzle Cross Sectional Area (ft <sup>2</sup> )	0.000885	0.00191	0.00341	0.00551	0.00767	0.01042	0.01364

\* meter box leak check = less than 0.02 cfm at 15 in. Hg or greatest vacuum observed during test.  
 † pilot tube leak check = hold 3 in. H<sub>2</sub>O for at least 15 sec.  
 ΔH = ΔP x K factor x [(T<sub>gim</sub> + 460)/528]

Point	Clock	Meter Reading	K	ΔP	ΔH	Desired ΔH	Vacuum	DGMT	Exit T	Filter T	Probe T	Stack T
PORT A												
1-A	8:26	644.322	6.0	0.15	0.90	0.90	0.5	66	50	72	---	74
1-B	8:36	651.200	6.0	0.15	0.92	0.90	0.5	73	41	72	---	73
1-C	8:46	657.610	6.0	0.15	0.92	0.90	0.5	77	40	74	---	74
			6.00	0.150	0.913	0.900	0.50	72.0	43.7	72.7	---	73.7
2-A	8:56	664.300	6.0	0.25	1.50	1.50	1.5	81	40	76	---	73
2-B	9:06	672.750	6.0	0.25	1.50	1.50	1.5	83	39	77	---	73
2-C	9:16	681.320	6.0	0.25	1.60	1.50	1.5	86	40	78	---	73
			6.00	0.250	1.533	1.500	1.50	83.3	39.7	77.0	---	73.0
3-A	9:26	689.830	6.0	0.31	1.90	1.86	2.0	88	42	79	---	74
3-B	9:36	699.700	5.0	0.31	1.90	1.55	2.0	89	42	79	---	74
3-C	9:46	708.920	5.0	0.31	1.90	1.55	2.0	90	42	78	---	74
			5.33	0.310	1.900	1.663	2.00	89.0	42.0	78.7	---	74.0
4-A	9:56	718.780	5.0	0.30	1.80	1.50	2.0	91	43	78	---	74
4-B	10:06	727.860	5.0	0.31	1.50	1.55	1.5	92	43	79	---	74
4-C	10:16	736.980	5.0	0.30	1.50	1.50	1.5	91	43	79	---	74
			5.00	0.303	1.600	1.517	1.67	91.3	43.0	78.7	---	74.0
PORT B												
5-A	10:27	745.631	5.0	0.14	0.70	0.70	0.5	92	45	79	---	76
5-B	10:37	751.460	5.0	0.14	0.70	0.70	0.5	92	41	79	---	77
5-C	10:47	757.310	5.0	0.14	0.70	0.70	0.5	92	40	80	---	77
			5.00	0.140	0.700	0.700	0.50	92.0	42.0	79.3	---	76.7
6-A	10:57	763.160	4.0	0.25	1.00	1.00	1.0	93	40	80	---	76
6-B	11:07	770.250	4.0	0.24	0.93	0.96	1.0	93	41	80	---	76
6-C	11:17	776.960	3.0	0.24	0.73	0.72	1.0	93	41	80	---	76
			3.67	0.243	0.887	0.893	1.00	93.0	40.7	80.0	---	76.0
7-A	11:27	783.190	3.0	0.30	0.90	0.90	0.5	94	42	80	---	76
7-B	11:37	789.680	3.0	0.30	0.90	0.90	0.7	94	42	80	---	76
7-C	11:47	796.320	3.0	0.31	0.90	0.93	0.7	95	44	80	---	76
			3.00	0.30	0.900	0.910	0.63	94.3	42.7	80.0	---	76.0
8-A	11:57	802.940	3.0	0.29	0.83	0.87	0.6	94	44	80	---	76
8-B	12:07	809.280	3.0	0.29	0.82	0.87	0.6	95	44	81	---	75
8-C	12:17	815.600	3.0	0.30	0.90	0.90	0.7	95	45	81	---	76
Final	12:27	822.268	3.00	0.29	0.850	0.880	0.63	94.7	44.3	80.7	---	75.7

# Isokinetic Worksheet

## TA 35-213 - Lathe #1 and Lathe #2, Run 2

Date: 8/20/24  
 Client: LANL  
 Plant: TA-35-213  
 Emission Source: Lathe #1 and Lathe #2  
 Technicians: CS, MM  
 DGM No: Console 2  
 DGM Factor, Y: 0.993  
 AH@: 1.624  
 Stack Diameter ("): 12  
 Stack Area (ft<sup>2</sup>): 0.785

Critical Isokinetic Test Data			Pre-Test Calibration Check		
Atmosph. Press. ("Hg)	22.88	Static P	V1	635.094	cf
Pitot Tube Factor, Cp	0.84	B	Vf	643.901	cf
Static Pressure (in H2O), P <sub>s</sub>	-0.16	-0.17	Vt	8.807	cf
M <sub>s</sub> - Molecular Weight (lb/lb-mole)	28.71	A	ΔH	1.60	"H2O
Dry Gas Fraction, 1-Ws	0.9881	-0.15	Time	10	min
Nozzle Diameter (in)	0.267	Nozzle (in.)	Im	540	R
Nozzle Area (ft <sup>2</sup> )	3.88E-04	0.26	Phar	22.88	"Hg
Total Sample Time (min)	240	0.27	constant	0.0319	"Hg-clm/R
Number of Points	8	0.27	Yc	0.9852	
Filter #	QF-17	0.267	ΔH@	1.5750	

Dry Gas Meter Leak Checks				Pitot Tube Leak Check	
pre-test	0.003 cf @	12"		+/-	4.8/4.7
post-test	0.000 cf @	9"		+/-	---
Moisture Data					
Run 1	Imp 1	Imp 2	Imp 3	Imp 4	Total Wt
Beginning	708.3	703.4	572.2	929.3	2913.2
Ending	707.9	705.2	574.2	962.4	2949.7
Total	-0.4	1.8	2.0	33.1	36.5
Gaseous Data (%)					
O2	20.90	CO2	0.04	N2	79.1
Scale Prelim Check (1000g): 1000.0 999.9 1000.0 999.97					

Start Vol	824.996	Stop Vol	1008.415	Actual Volume	183.419	Target	180.000	ACF x Y	182.135	SCF	134.872	
Start Time	13:00	Stop Time	10:07	Sample Time per Point	30	K Factor	5.0	Z ΔP	4.058	Avg ΔP	0.262	
											Isokinetic Rate (%):	98.71

Sample Point per Port	Start Meter Time	Observed Clock Time	Observed Meter Reading (ft <sup>3</sup> )	Next Point Target (ft <sup>3</sup> )	ΔP ("H2O)	ΔH ("H2O)	Desired ΔH ("H2O)	Observed K Factor	Sample Vacuum ("Hg)	DGM Temp. (°F)	Last Imp. (°F)	Filter Temp. (°F)	Probe Temp. (°F)	Stack Temp. (°F)	Point Vol Sampled (DSCF)	Σ - ΔP ("H2O)	Stack Velocity (ft/s)	Average ISO (%)	
1	0:00	13:00	824.996	843.537	0.160	0.733	0.80	4.37	0.40	94.0	59.7	85.0	---	78.3	12.814	0.400	26.01	95.3	
2	30:00		842.660	866.189	0.257	1.175	1.28	4.36	0.47	95.0	47.3	79.3	---	76.7	16.171	0.507	32.89	95.0	
3	60:00		864.960	889.732	0.283	1.273	1.42	4.26	0.50	96.7	49.0	77.7	---	77.3	17.104	0.532	34.58	95.2	
4	90:00		888.610	913.177	0.277	1.257	1.38	4.29	0.50	99.0	53.0	78.3	---	78.0	16.934	0.526	34.19	95.3	
5	120:00	8:07	912.125	930.051	0.163	0.733	0.82	4.49	0.40	68.0	47.0	66.3	---	74.0	13.195	0.404	26.17	95.5	
6	150:00		929.460	952.838	0.270	1.383	1.35	5.04	0.70	76.7	45.3	70.0	---	74.0	18.243	0.520	33.65	97.1	
7	180:00		953.770	980.295	0.340	1.700	1.70	4.87	1.00	82.3	47.0	71.3	---	72.3	20.375	0.583	37.70	98.1	
8	210:00		981.180	1008.078	0.343	1.717	1.72	4.82	1.00	88.0	48.0	72.3	---	73.7	20.036	0.586	37.93	98.6	
Finish	240:00	10:07	1008.415	---	---	---	---	---	---	---	---	---	---	---	Sum	Sum	---	---	---
Total	Total	Total	183.419	Average	0.262	1.246	1.308	4.562	0.621	87.458	49.542	75.042	---	75.542	134.8719	4.0577	32.890	96.27	

BG Nozzle Number	1	2	3	4	5	6	7
Nozzle Diameter (in.)	0.125	0.187	0.250	0.312	0.375	0.437	0.500
Nozzle Cross Sectional Area (ft <sup>2</sup> )	0.000085	0.000191	0.000341	0.000531	0.000767	0.001042	0.001364

\* meter box leak check = less than 0.02 cfm at 15 in. Hg or greatest vacuum observed during test.  
 † pitot tube leak check = hold 3 in. H<sub>2</sub>O for at least 15 sec.  
 ΔH = ΔP x K factor x [(T<sub>gim</sub> + 460)/528]

Point	Clock	Meter Reading	K	ΔP	ΔH	Desired AH	Vacuum	DGMT	Exit T	Filter T	Probe T	Stack T
PORT A												
1-A	13:00	824.996	4.5	0.15	0.70	0.68	0.4	94	78	89	---	80
1-B	13:10	830.670	4.5	0.17	0.77	0.77	0.4	94	53	84	---	80
1-C	13:20	836.765	4.5	0.16	0.73	0.72	0.4	94	48	82	---	77
			4.50	0.160	0.733	0.720	0.40	94.0	59.7	85.0	---	78.3
2-A	13:30	842.660	4.5	0.25	1.13	1.13	0.4	94	47	81	---	76
2-B	13:40	849.810	4.5	0.26	1.20	1.17	0.5	96	47	79	---	77
2-C	13:50	857.300	4.5	0.26	1.20	1.17	0.5	95	48	78	---	77
			4.50	0.257	1.175	1.155	0.47	95.0	47.3	79.3	---	76.7
3-A	14:00	864.960	4.5	0.29	1.30	1.31	0.5	96	48	77	---	77
3-B	14:10	872.715	4.5	0.28	1.26	1.26	0.5	96	48	77	---	77
3-C	14:20	880.590	4.5	0.28	1.26	1.26	0.5	98	51	79	---	78
			4.50	0.283	1.273	1.275	0.50	96.7	49.0	77.7	---	77.3
4-A	14:30	888.610	4.5	0.28	1.26	1.26	0.5	98	52	79	---	78
4-B	14:40	896.335	4.5	0.27	1.25	1.22	0.5	100	53	79	---	78
4-C	14:50	904.225	4.5	0.28	1.26	1.26	0.5	99	54	77	---	78
			4.50	0.277	1.257	1.245	0.50	99.0	53.0	78.3	---	78.0
PORT B												
5-A	8:07	912.125	4.5	0.16	0.72	0.72	0.4	65	52	63	---	73
5-B	8:17	917.910	4.5	0.16	0.72	0.72	0.4	68	45	67	---	74
5-C	8:27	923.670	4.5	0.17	0.76	0.77	0.4	71	44	69	---	75
			4.50	0.163	0.733	0.735	0.40	68.0	47.0	66.3	---	74.0
6-A	8:37	929.460	5.0	0.27	1.35	1.35	0.6	74	43	69	---	74
6-B	8:47	937.370	5.0	0.27	1.40	1.35	0.7	77	46	70	---	74
6-C	8:57	945.560	5.0	0.27	1.40	1.35	0.8	79	47	71	---	74
			5.00	0.270	1.383	1.350	0.70	76.7	45.3	70.0	---	74.0
7-A	9:07	953.770	5.0	0.34	1.70	1.70	1.0	81	47	70	---	74
7-B	9:17	963.100	5.0	0.34	1.70	1.70	1.0	82	47	73	---	70
7-C	9:27	971.950	5.0	0.34	1.70	1.70	1.0	84	47	71	---	73
			5.00	0.34	1.700	1.700	1.00	82.3	47.0	71.3	---	72.3
8-A	9:37	981.180	5.0	0.35	1.75	1.75	1.0	87	48	72	---	74
8-B	9:47	990.280	5.0	0.34	1.70	1.70	1.0	88	48	72	---	73
8-C	9:57	999.340	5.0	0.34	1.70	1.70	1.0	89	48	73	---	74
Final	10:07	1008.415	5.00	0.34	1.717	1.717	1.00	88.0	48.0	72.3	---	73.7

# Isokinetic Worksheet

## TA 35-213 - Lathe #1 and Lathe #2, Run 3

Date: 8/21/24  
 Client: LANL  
 Plant: TA-35-213  
 Emission Source: Lathe #1 and Lathe #2  
 Technicians: CS, MM  
 DGM No: Console 2  
 DGM Factor, Y: 0.993  
 ΔH@: 1.624  
 Stack Diameter ("): 12  
 Stack Area (ft²): 0.785

Critical Isokinetic Test Data			Pre-Test Calibration Check		
Atmosph. Press. ("Hg)	22.88	Static P	V1	635.094	cf
Pilot Tube Factor, Cp	0.84	B	Vf	643.901	cf
Static Pressure (in H2O), P <sub>s</sub>	-0.16	-0.17	Vt	8.807	cf
M <sub>s</sub> - Molecular Weight (lb/lb-mole)	28.71	A	ΔH	1.60	"H2O
Dry Gas Fraction, 1-Bws	0.9881	-0.15	Time	10	min
Nozzle Diameter (in)	0.267	Nozzle (in.)	Im	540	R
Nozzle Area (ft²)	3.88E-04	0.26	Phar	22.88	"Hg
Total Sample Time (min)	240	0.27	constant	0.0319	"Hg-clm/R
Number of Points	8	0.27	Yc	0.9852	
Filter #	QF-18	0.267	ΔH@	1.5750	

Dry Gas Meter Leak Checks				Pilot Tube Leak Check	
pre-test	0.002 cf @	14"		+/-	4.8/4.7
post-test	0.001 cf @	8"		+/-	---
Moisture Data					
Run 1	Imp 1	Imp 2	Imp 3	Imp 4	Total Wt
Beginning	715.1	689.5	597.0	949.1	2950.7
Ending	717.6	691.9	597.8	982.7	2990.0
Total	2.5	2.4	0.8	33.6	39.3
Gaseous Data (%)					
O2	20.90	CO2	0.04	N2	79.1
Scale Prelim Check (1000g): 1000.0 999.9 1000.0 999.97					

Start Vol	1008.710	Stop Vol	1199.971	Actual Volume	191.261	Target	180.000	ACF x Y	189.922	SCF	139.075
Start Time	10:27	Stop Time	12:59	Sample Time per Point	30	K Factor	5.0	Σ √ΔP	4.091	Avg √ΔP	0.265

Isokinetic Rate (%):	101.12
----------------------	--------

Sample Point per Port	Start Meter Time	Observed Clock Time	Observed Meter Reading (ft³)	Next Point Target (ft³)	ΔP ("H2O)	ΔH ("H2O)	Desired ΔH ("H2O)	Observed K Factor	Sample Vacuum ("Hg)	DGM Temp. (°F)	Last Imp. (°F)	Filter Temp. (°F)	Probe Temp. (°F)	Stack Temp. (°F)	Point Vol Sampled (DSCF)	Σ - √ΔP ("H2O)	Stack Velocity (ft/s)	Average ISO (%)	
1	0:00	10:27	1008.710	1027.694	0.170	0.850	0.85	4.80	0.50	89.7	48.3	77.3	---	76.7	14.043	0.412	26.77	101.1	
2	30:00		1027.910	1051.496	0.263	1.317	1.32	4.81	0.67	89.3	43.7	77.3	---	76.3	17.419	0.513	33.30	100.9	
3	60:00		1051.675	1077.904	0.323	1.617	1.62	4.79	1.13	91.7	45.3	78.7	---	76.0	19.382	0.569	36.89	101.0	
4	90:00		1078.205	1105.062	0.337	1.683	1.68	4.77	1.37	93.0	47.3	80.7	---	74.7	19.747	0.580	37.60	101.0	
5	120:00	12:29	1105.295	1124.388	0.170	0.850	0.85	4.77	0.63	94.0	49.0	81.3	---	79.0	14.082	0.412	26.83	101.1	
6	150:00		1124.700	1148.094	0.253	1.300	1.27	4.87	1.00	96.0	55.3	81.7	---	77.3	17.263	0.503	32.70	101.2	
7	180:00		1148.540	1174.035	0.300	1.500	1.50	4.74	1.00	97.0	49.7	77.3	---	77.0	18.560	0.548	35.57	101.2	
8	210:00		1174.200	1200.039	0.307	1.567	1.53	4.83	1.03	99.0	50.7	79.3	---	78.0	18.577	0.554	36.00	101.0	
Finish	240:00	12:59	1199.971	---	---	---	---	---	---	---	---	---	---	---	Sum	Sum	---	---	---
Total	Total		191.261	Average	0.265	1.335	1.327	4.796	0.917	93.708	48.667	79.208	---	76.875	139.0746	4.0915	33.206	101.07	

BG Nozzle Number	1	2	3	4	5	6	7
Nozzle Diameter (in.)	0.125	0.187	0.250	0.312	0.375	0.437	0.500
Nozzle Cross Sectional Area (ft²)	0.000885	0.00191	0.00341	0.00531	0.00767	0.01042	0.01364

\* meter box leak check = less than 0.02 cfm at 15 in. Hg or greatest vacuum observed during test.  
 † pilot tube leak check = hold 3 in. H<sub>2</sub>O for at least 15 sec.  
 ΔH = ΔP x K factor x [(T<sub>gim</sub> + 460)/528]

Point	Clock	Meter Reading	K	ΔP	ΔH	Desired ΔH	Vacuum	DGMT	Exit T	Filter T	Probe T	Stack T
PORT A												
1-A	10:27	1008.710	5.0	0.17	0.85	0.85	0.5	90	56	78	---	76
1-B	10:37	1015.150	5.0	0.17	0.85	0.85	0.5	90	45	77	---	77
1-C	10:47	1021.515	5.0	0.17	0.85	0.85	0.5	89	44	77	---	77
			5.00	0.170	0.850	0.850	0.50	89.7	48.3	77.3	---	76.7
2-A	10:57	1027.910	5.0	0.25	1.25	1.25	0.5	89	44	77	---	77
2-B	11:07	1035.550	5.0	0.27	1.35	1.35	0.5	89	43	77	---	76
2-C	11:17	1043.600	5.0	0.27	1.35	1.35	1.0	90	44	78	---	76
			5.00	0.263	1.317	1.317	0.67	89.3	43.7	77.3	---	76.3
3-A	11:27	1051.675	5.0	0.32	1.60	1.60	1.0	91	45	79	---	76
3-B	11:37	1060.420	5.0	0.33	1.65	1.65	1.2	92	45	79	---	76
3-C	11:47	1069.420	5.0	0.32	1.60	1.60	1.2	92	46	78	---	76
			5.00	0.323	1.617	1.617	1.13	91.7	45.3	78.7	---	76.0
4-A	11:57	1078.205	5.0	0.34	1.70	1.70	1.4	93	47	80	---	76
4-B	12:07	1087.270	5.0	0.33	1.65	1.65	1.4	93	47	81	---	72
4-C	12:17	1096.250	5.0	0.34	1.70	1.70	1.3	93	48	81	---	76
			5.00	0.337	1.683	1.683	1.37	93.0	47.3	80.7	---	74.7
PORT B												
5-A	12:29	1105.295	5.0	0.17	0.85	0.85	0.5	94	48	82	---	79
5-B	12:39	1110.730	5.0	0.17	0.85	0.85	0.7	94	49	81	---	79
5-C	12:49	1116.170	5.0	0.17	0.85	0.85	0.7	94	50	81	---	79
			5.00	0.170	0.850	0.850	0.63	94.0	49.0	81.3	---	79.0
6-A	12:59	1124.700	5.0	0.26	1.30	1.30	1.0	95	52	82	---	77
6-B	13:09	1132.685	5.0	0.25	1.30	1.25	1.0	96	56	83	---	77
6-C	13:19	1140.600	5.0	0.25	1.30	1.25	1.0	97	58	80	---	78
			5.00	0.253	1.300	1.267	1.00	96.0	55.3	81.7	---	77.3
7-A	13:29	1148.540	5.0	0.30	1.50	1.50	1.0	97	51	78	---	77
7-B	13:39	1157.030	5.0	0.30	1.50	1.50	1.0	97	49	78	---	77
7-C	13:49	1166.000	5.0	0.30	1.50	1.50	1.0	97	49	76	---	77
			5.00	0.30	1.500	1.500	1.00	97.0	49.7	77.3	---	77.0
8-A	13:59	1174.200	5.0	0.30	1.50	1.50	1.0	98	50	77	---	78
8-B	14:09	1182.860	5.0	0.31	1.60	1.55	1.1	99	50	80	---	78
8-C	14:19	1191.525	5.0	0.31	1.60	1.55	1.0	100	52	81	---	78
Final	14:29	1200.416	5.00	0.31	1.567	1.533	1.03	99.0	50.7	79.3	---	78.0

## **Appendix C: Example Calculations**

## Example Calculations

### Moisture Content and Stack Molecular Weight

Moisture Content Determination			
<i>Moisture Content via EPA Method 4 - Run 1</i>			
V <sub>1</sub> =	Initial dry gas meter reading	=	644.322 ft <sup>3</sup>
V <sub>2</sub> =	Final dry gas meter reading	=	822.268 ft <sup>3</sup>
V <sub>M</sub> =	Total cubic meters of stack gas metered (V <sub>2</sub> -V <sub>1</sub> )	=	<b>177.946 ft<sup>3</sup></b>
Y =	Dry gas meter correction factor	=	0.9930 unitless
V <sub>M (corr)</sub> =	Corrected DGM volume to cubic feet (V <sub>M</sub> x Y)	=	<b>176.700 ft<sup>3</sup></b>
W <sub>I</sub> =	Initial weight of impinger train	=	2971.4 grams
W <sub>F</sub> =	Final weight of impinger train	=	3004.5 grams
W <sub>TOT</sub> =	Total weight gain of impinger train (W <sub>F</sub> -W <sub>I</sub> )	=	<b>33.1 grams</b>
K <sub>1</sub> =	Conversion factor	=	1.333 L <sub>H2O</sub>
1 m <sup>3</sup> /mL =	Conversion factor	=	1 x e 6
D <sub>H2O</sub> =	Density of water	=	998.2 L/g
K <sub>1A</sub> =	Conv. factor x density of water @ EPA STP	=	1.336 L/g
K <sub>3</sub> =	$\frac{528^{\circ}\text{R} \times 28.316 \text{ L} / \text{ft}^3}{29.92 \text{ " Hg}}$ @ EPA STP	=	499.7 $\frac{^{\circ}\text{R} - \text{L}}{\text{"Hg-ft}^3}$
P <sub>ATM</sub> =	Atmospheric Pressure	=	<b>22.88 "Hg</b>
T <sub>°F</sub> =	Average temperature of dry gas meter	=	88.71 °F
T <sub>°R</sub> =	Avg. temperature of dry gas meter + 460°	=	<b>548.71 °R</b>
<i>Formulas and Calculations (Equation 4-4)</i>			
V <sub>M (std)</sub> =	Volume of gas metered @ EPA STP	=	
V <sub>M (std)</sub> =	$\frac{176.700 \times 22.88 \times 499.7}{548.71}$	@ EPA STP =	<b>3681.938 L</b>
B <sub>WS</sub> =	Moisture content by volume	=	
B <sub>WS</sub> =	$\frac{(33.10 \times 1.336)}{(33.10 \times 1.336) + 3681.94}$	=	0.011866
		x 100 =	<b>1.19% Moisture</b>
<i>Stack Gas Molecular Weight</i>			
MW <sub>H2O</sub> =	Molecular weight of water	=	18 lb/lb-mol
MW <sub>O2</sub> =	Molecular weight of oxygen	=	32 lb/lb-mol
MW <sub>CO2</sub> =	Molecular weight of carbon dioxide	=	44 lb/lb-mol
MW <sub>N2</sub> =	Molecular weight of nitrogen	=	28 lb/lb-mol
C <sub>O2</sub> =	Volume fraction of corrected oxygen	=	0.2090 O <sub>2</sub>
C <sub>CO2</sub> =	Volume fraction of corrected carbon dioxide	=	0.0004 CO <sub>2</sub>
C <sub>N2</sub> =	Volume fraction of nitrogen = 1-(C <sub>O2</sub> + C <sub>CO2</sub> )	=	0.7906 N <sub>2</sub>
1-B <sub>WS</sub> =	Dry gas fraction = (1 - B <sub>WS</sub> )	=	0.9881 Dry Exhaust
<i>Formulas and Calculations (Equations 3-1 and 2-5)</i>			
M <sub>s</sub> =	Wet molecular weight of stack gas	=	lb/lb-mol
M <sub>D</sub> =	Dry molecular weight of stack gas	=	lb/lb-mol
M <sub>D</sub> =	$(\text{MW}_{\text{O}_2} \times \text{C}_{\text{O}_2}) + (\text{MW}_{\text{CO}_2} \times \text{C}_{\text{CO}_2}) + (\text{MW}_{\text{N}_2} \times \text{C}_{\text{N}_2})$	=	
	$= 6.688 + 0.018 + 22.137$	=	<b>28.842 lb/lb-mol</b>
M <sub>s</sub> =	$(18 \times \text{B}_{\text{WS}}) + [(1-\text{B}_{\text{WS}}) \times \text{M}_{\text{D}}]$	=	
	$= 0.214 + 28.500$	=	<b>28.714 lb/lb-mol</b>

## Example Calculations Exhaust Flow Rates

### Stack Gas Velocity Determination

*Velocity and stack flow rate via Pitot tube (Equations 2-6, 2-8, 2-9, 2-10)*

$C_P$	S-Type Pitot Tube Coefficient (dimensionless)	=	<b>0.84</b>
$\Delta P$	Differential Pressure Measured by Pitot Tube	=	"H <sub>2</sub> O
$\sqrt{\Delta P_{avg}}$	Average Square Root of $\Delta P$ 's	=	<b>0.494</b> "H <sub>2</sub> O
$T_{S-^{\circ}F}$	Average Temperature (measured)	=	74.9 °F
$T_{S-^{\circ}R}$	Absolute Temperature (°R) = (Ts + 460)	=	<b>534.9</b> °R
1-BWS	Dry Gas Fraction (unitless)	=	0.9881
$M_S$	Molecular Weight of Stack Gas, wet	=	28.71 lbs/lbs-mol
$K_P$	Pitot Tube Constant in Minutes	=	<b>5129.4</b> ft/min
$K_Y$	Standard Pressure/Temperature Coefficient	=	°R-min/"Hg-hr
	= 528° R ÷ 29.92 "Hg x 60 min ÷ 1 hr	=	1058.8 $\frac{^{\circ}R-min}{"Hg-hr}$
$P_B$	Atmospheric Pressure ("Hg)	=	22.88 "Hg
$P_G$	Static Pressure ("H <sub>2</sub> O)	=	-0.16 "H <sub>2</sub> O
$P_S$	Absolute Pressure	=	"Hg
$P_S$	$P_B + \frac{P_G}{13.6 ("H_2O/"Hg)}$	=	<b>22.87</b> "Hg
$D$	Stack Diameter	=	1.00 feet
$A_S$	Area of Stack (ft <sup>2</sup> ) = $\pi \times d_i^2 \div 4$	=	<b>0.79</b> ft <sup>2</sup>
$V_S$	Stack Velocity	=	ft/min
	= $K_P \times C_P \times \sqrt{\Delta P_{avg}} \times$	=	ft/min
$V_S$	= 4308.7 x 0.49 x $\frac{535}{656.64}$	=	1922.32 ft/min
$Q_A$	Volumetric Flow Rate = $V_S \times A_S$	=	<b>1,509.78</b> ft <sup>3</sup> /min
$Q_D$	Stack flow rate on a dry basis and standard conditions		
$Q_{STD}$	Volumetric Flow Rate on a Dry Basis		
	= $Q_A \times (1-BWS) \times K_Y \times P_S \div T_S$		
$Q_{STI}$	= 1,509.78 x 1046.3 x $\frac{22.87}{534.9}$	=	<b>67,537</b> DSCF/H

## Example Calculations Concentration and Mass Emission Rates

<b>Calculated Emissions</b>			
<i>Determination of Be Concentrations and Mass Emission Rates</i>			
Q <sub>S M1-4</sub> =	Measured Stack Flow Rate	=	67,537 SCF/H Dry
V <sub>M-DRY</sub> =	Dry Standard Metered Volume	=	130.593 SCF Dry
Be Mass =	Mass of Be Collected (Blank Corrected)	=	0.20 μg
CF <sub>2</sub> =	Conversion from micrograms to grams	=	1.00E-06 g / μg
CF <sub>3</sub> =	Conversion from grains to pounds	=	453.6 gr/lb
CF <sub>4</sub> =	Conversion from sample time to hr	=	0.25 hr
<p><b>Concentration (g/DSCF) = <math>\frac{\text{mass collected } (\mu\text{g}) \times \text{CF}_2}{\text{V}_{\text{M-DRY}} (\text{SCF})}</math></b></p> <p><b>Be CONC = <math>\frac{0.20 \times 1.00\text{E-}06}{130.593} = 1.53\text{E-}09 \text{ gr/DSCF}</math></b></p>			
<p><b>Mass Emissions (lbs/hr) = TSP<sub>CONC</sub> ÷ CF<sub>3</sub> x CF<sub>4</sub> x Q<sub>S M1-4</sub></b></p> <p><b>BE g/hr = 1.53E-09 x 67,537 x 0.25 = 2.59E-05 g/hr</b></p>			
<i>Determination of Isokinetics Conditions</i>			
V <sub>M-DRY</sub> =	Dry Standard Metered Volume	=	130.593 SCF Dry
θ =	Total Sampling Time	=	240 min
K <sub>5</sub> =	Atmospheric Sampling Standard	=	0.0945 "Hg-min/°R-sec
V <sub>S</sub> =	Stack Gas Velocity	=	32.04 ft/sec
T <sub>S</sub> =	Absolute Stack Temperature	=	534.88 °R-sec
P <sub>S</sub> =	Absolute Stack Pressure	=	22.87 "Hg
A <sub>N</sub> =	Area of Nozzle Diameter	=	3.88E-04 ft <sup>2</sup>
1-B <sub>WS</sub> =	Dry Gas Fraction (unitless)	=	0.9881
<p><b>Isokinetic Variation (%) = <math>\frac{\text{K}_5 \times \text{T}_S \times \text{V}_{\text{m-std}}}{\text{P}_S \times \text{V}_S \times \text{A}_N \times \theta \times (1-\text{B}_{\text{WS}})}</math></b></p> <p><b>I (%) = <math>\frac{6600.90}{67.391} = 97.95 \%</math></b></p>			



**Appendix D:  
QA/QC Activities and Equipment Calibration Results**

**APEX INSTRUMENTS METHOD 5 PRE-TEST CONSOLE CALIBRATION**  
**USING CALIBRATED CRITICAL ORIFICES**  
**5-POINT ENGLISH UNITS**

Meter Console Information	
Console Model Number	MC522 #2
Console Serial Number	1205014
DGM Model Number	MS-4
DGM Serial Number	30-Nov-24

Calibration Conditions			
Date	Time	8/5/24	14:00:00
Barometric Pressure		25.0	in Hg
Theoretical Critical Vacuum <sup>1</sup>		11.8	in Hg
Calibration Technician		SV	

Factors/Conversions	
Std Temp	528
Std Press	29.92
K <sub>1</sub>	17.647

<sup>1</sup>For valid test results, the Actual Vacuum should be 1 to 2 in. Hg greater than the Theoretical Critical Vacuum shown above.

<sup>2</sup>The Critical Orifice Coefficient, K', must be entered in English units, (ft<sup>3</sup>•°R<sup>1/2</sup>)/(in.Hg•min).

Calibration Data										
Run Time	Metering Console					Critical Orifice				
Elapsed	DGM Orifice ΔH	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final	Serial Number	Coefficient	Amb Temp Initial	Amb Temp Final	Actual Vacuum
(θ)	(P <sub>m</sub> )	(V <sub>m</sub> )	(V <sub>mf</sub> )	(t <sub>mi</sub> )	(t <sub>mf</sub> )		K'	(t <sub>amb</sub> )	(t <sub>amb</sub> )	
min	in H <sub>2</sub> O	cubic feet	cubic feet	°F	°F		see above <sup>2</sup>	°F	°F	in Hg
15.0	0.28	615.364	620.332	84	85	TC 40	0.2536	86	86	21
11.0	0.56	620.332	625.735	85	85	TC 48	0.3690	86	86	20
8.0	0.97	625.735	630.985	85	85	TC 55	0.4930	86	88	19
6.0	1.70	630.985	636.213	85	86	TC 63	0.6550	88	88	17
5.0	2.80	636.213	641.924	86	86	TC 73	0.8580	88	88	16

Results								
Standardized Data				Dry Gas Meter				
Dry Gas Meter		Critical Orifice		Calibration Factor		Flowrate	ΔH @	
(V <sub>m(std)</sub> )	(Q <sub>m(std)</sub> )	(V <sub>Cr(std)</sub> )	(Q <sub>Cr(std)</sub> )	Value	Variation	Std & Corr	0.75 SCFM	Variation
(Y)	(ΔY)	(Q <sub>m(std)(corr)</sub> )	(ΔH@)	(ΔΔH@)				
cubic feet	cfm	cubic feet	cfm			cfm	in H <sub>2</sub> O	
4.029	0.269	4.070	0.271	1.010	0.018	0.271	1.738	0.113
4.381	0.398	4.343	0.395	0.991	-0.001	0.395	1.643	0.019
4.262	0.533	4.216	0.527	0.989	-0.003	0.527	1.601	-0.023
4.249	0.708	4.197	0.700	0.988	-0.005	0.700	1.598	-0.027
4.653	0.931	4.581	0.916	0.985	-0.008	0.916	1.542	-0.082
				0.993	Y Average		1.624	ΔH@ Average

Note: For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is ±0.02.

I certify that the above Dry Gas Meter was calibrated in accordance with USEPA Methods, CFR Title 40, Part 60, Appendix A-3, Method 5, 16.2.3

Signature

Date

**APEX INSTRUMENTS METHOD 5 POST-TEST CONSOLE CALIBRATION**  
**USING CALIBRATED CRITICAL ORIFICES**  
**3-POINT ENGLISH UNITS**

Meter Console Information	
Console Model Number	MC522 #2
Console Serial Number	1205014
DGM Model Number	MS-4
DGM Serial Number	1506596

Calibration Conditions			
Date	Time	8/22/24	10:52
Barometric Pressure		25.0	in Hg
Theoretical Critical Vacuum <sup>1</sup>		11.8	in Hg
Calibration Technician		MM	

Factors/Conversions		
Std Temp	528	°R
Std Press	29.92	in Hg
K <sub>c</sub>	17.647	oR/in Hg

<sup>1</sup>For valid test results, the Actual Vacuum should be 1 to 2 in. Hg greater than the Theoretical Critical Vacuum shown above.

<sup>2</sup>The Critical Orifice Coefficient, K', must be entered in English units, (ft<sup>3</sup>°R<sup>1/2</sup>)/(in.Hg\*min).

Calibration Data										
Run Time	Metering Console					Critical Orifice				
Elapsed	DGM Orifice ΔH	Volume Initial	Volume Final	Outlet Temp Initial	Outlet Temp Final	Serial Number	Coefficient	Amb Temp Initial	Amb Temp Final	Actual Vacuum
(θ)	(P <sub>m</sub> )	(V <sub>mi</sub> )	(V <sub>mf</sub> )	(t <sub>mi</sub> )	(t <sub>mf</sub> )		K'	(t <sub>amb</sub> )	(t <sub>amb</sub> )	
min	in H <sub>2</sub> O	cubic feet	cubic feet	°F	°F		see above <sup>2</sup>	°F	°F	in Hg
8.0	0.970	1200.579	1205.787	80	81	55	0.493	83	83	19.0
8.0	0.970	1205.787	1211.018	81	81	55	0.493	83	83	19.0
8.0	0.970	1211.018	1216.238	81	81	55	0.493	83	83	19.0

Results								
Standardized Data				Dry Gas Meter				
Dry Gas Meter		Critical Orifice		Calibration Factor		Flowrate	ΔH @	
(V <sub>m(std)</sub> )	(Q <sub>m(std)</sub> )	(V <sub>cr(std)</sub> )	(Q <sub>cr(std)</sub> )	Value	Variation	Std & Corr	0.75 SCFM	Variation
cubic feet	cfm	cubic feet	cfm	(Y)	(ΔY)	(Q <sub>m(std)(corr)</sub> )	(ΔH@)	(ΔΔH@)
						cfm	in H <sub>2</sub> O	
4.263	0.535	4.217	0.529	0.989	-0.001	0.529	1.601	0.001
4.278	0.535	4.233	0.529	0.990	-0.001	0.529	1.599	0.000
4.269	0.534	4.233	0.529	0.992	0.001	0.529	1.599	0.000
<b>Pretest Gamma</b>	0.989	<b>% Deviation</b>	0.117	0.990	Y Average		1.600	ΔH@ Average


Note: For Calibration Factor Y, the ratio of the reading of the calibration meter to the dry gas meter, acceptable tolerance of individual values from the average is +/-0.02.

I certify that the above Dry Gas Meter was calibrated in accordance with USEPA Methods, CFR Title 40, Part 60, Appendix A-3, Method 5, 16.2.3


Signature

Date

# Meter Gamma vs Flowrate



### Meter Pressure vs Flowrate



## Metering Console Thermocouple Calibration

**Date:** 8/22/24  
**Technician:** MM  
**Meter ID:** Apex XC522 Console #2

**Thermocouple ID:** Alltec Series 22 TC Source #8766041  
**Thermocouple:** PT-8-1  
**Ambient Temp.:** 82.5

STANDARD TEMP. (°R)	AUX		STACK		PROBE		OVEN		FILTER		EXIT		AVERAGE DIFF. (%)
	Rdg °R	Diff.%	Rdg °R	Diff.%	Rdg °R	Diff.%	Rdg °R	Diff.%	Rdg °R	Diff.%	Rdg °R	Diff.%	
560	555	0.01	556	0.01	555	0.01	557	0.01	555	0.01	556	0.01	0.01
760	757	0.00	758	0.00	757	0.00	755	0.01	757	0.00	757	0.00	0.00
960	954	0.01	955	0.01	954	0.01	952	0.01	953	0.01	955	0.01	0.01
1160	1156	0.00	1156	0.00	1156	0.00	1157	0.00	1153	0.01	1157	0.00	0.00
1360	1356	0.00	1355	0.00	1357	0.00	1355	0.00	1356	0.00	1357	0.00	0.00
1560	1556	0.00	1555	0.00	1557	0.00	1555	0.00	1556	0.00	1557	0.00	0.00
1760	1755	0.00	1746	0.01	1756	0.00	1754	0.00	1751	0.01	1756	0.00	0.00
1960	1947	0.01	1955	0.00	1956	0.00	1954	0.00	1955	0.00	1956	0.00	0.00
2160	2153	0.00	2154	0.00	2156	0.00	2155	0.00	2154	0.00	2156	0.00	0.00
2360	2351	0.00	2350	0.00	2356	0.00	2349	0.00	2355	0.00	2355	0.00	0.00

Standard Temp. °R = °F + 460°

Average Percent Error = **0.00%**

Percent Error =  $\frac{\text{Standard Temp.} - \text{Measured Temp.}}{\text{Standard Temp.}} \times 100$

Tolerance ± **1.5%**



# Temperature Sensor Calibration Data Sheet

<i>Thermocouple Calibration</i>			
<i>Method: 2 Section 10.3.1</i>			
<i>Frequency: After each field use</i>			
<i>Criteria: &lt; 1.5 % Absolute Temperature</i>			
Altitude (ft): 5087		Calibration Date: 8/22/24	
Pressure ("Hg): 24.81		Technician: MM	
Boiling Point @ Alt.: 202.81		Thermocouple ID: PT 3-1	
<i>Water Temperature Range</i>	<i>Freezing</i>	<i>Room</i>	<i>Boiling</i>
Reference Thermometer (°F)		84.2	
Probe Thermometer (°F)		84.0	
Reference Thermometer (°R)	460.0	544.2	460.0
Probe Thermometer (°R)	460.0	544.0	460.0
<b>Absolute Temperature Difference</b>	<b>0.0%</b>	<b>0.04%</b>	<b>0.00%</b>
Absolute Temp. (°R) = 460+°F			
Alt. Correction for Water Boiling Point = 49.161*Ln(I"Hg)+44.932			
Pressure ("Hg) = 29.921*(1-6.8753e-6*alt)^5.2559			

# S-Type Pitot Tube Calibration

Date: 8/22/24  
 Technician: MM  
 Pitot Tube ID: PT-3-1 Assembly  
 Pitot Tube Type: Method 5 Assembly S-Type - 3'

- (a) end view; face opening planes perpendicular to transverse axis;
- (b) top view; face opening planes parallel to longitudinal axis;
- (c) side view; both legs of equal length and centerlines coincident, when viewed from both sides. Baseline coefficient values of 0.84 may be assigned to pitot tubes constructed this way.



Dt-a =	<u>0.375</u> "	<u>1.05Dt</u>	<u>Pa</u>	<u>1.5Dt</u>
Dt-b =	<u>0.375</u> "	0.3938	0.472	0.5625
Pa =	<u>0.472</u> "	<u>1.05Dt</u>	<u>Pb</u>	<u>1.5Dt</u>
Pb =	<u>0.472</u> "	0.3938	0.472	0.5625
P =	<u>0.944</u> "			

**Method 2 - 10.1:**  
 Coefficient = 0.84 if:  
 $0.188'' \leq Dt \leq 0.375''$  and if  $Pa=Pb$  and  $1.05Dt < Pa/2Dt < 1.50Dt$ .

The types of face-opening misalignment shown above will not affect the baseline value of  $C_{pit}$  so long as  $\alpha_1$  and  $\alpha_2 \leq 10^\circ$ ,  $\beta_1$  and  $\beta_2 \leq 5^\circ$ ,  $z \leq 0.32$  cm (1/8 in.), and  $w \leq 0.08$  cm (1/32 in.) (Reference 11.0 in Section 16.0)



$$\alpha-1 = \frac{0^\circ}{0^\circ}$$



$$\beta-1 = \frac{2^\circ}{2^\circ}$$



$$\beta-1 = \frac{0^\circ}{0^\circ}$$



$$z = \frac{0.00''}{z \leq}$$



$$Wt = \frac{0.00''}{Wt \leq}$$



**Appendix E:**  
**Laboratory Analysis Report and Chain of Custody**

COMPLIANCE SERVICES & TESTING

PROJECT: LANL TA-35

CLIENT # C062  
REPORT # 24-512

SUBMITTED BY:  
**Alliance Technical Group**  
12242 S.W. GARDEN PLACE  
TIGARD, OR 97223  
(503) 624-2183



12242 SW Garden Place ❖ Tigard, OR 97223-8246 ❖ (503) 624-2183

---

## **Case Narrative**

Date: September 5, 2024

### **General Information**

Client: Compliance Services & Testing  
Client Number: C062  
Report Number: 24-512  
Sample Description: Impinger Trains  
Sample Numbers: 24-S2462 – 24-S2470

### **Analysis**

Analytes: Be  
Analytical Protocols: EPA Method 29 (8/2/17 version)  
Analytical Notes: No problems were encountered during the analyses. The back samples consisted of impinger catches and nitric rinses. The results are not blank corrected.  
QA/QC Review: All the data have been reviewed by the analysts performing the analyses and the project manager. All the quality control and sample-specific information in this package is complete and meets or exceeds the minimum requirements for acceptability.  
Comments: If you have any questions or concerns regarding this analysis, please feel free to contact the project manager.  
Disclaimer: This report shall not be reproduced, except in full, without the written approval of the laboratory. The results only represent those of the samples as received into the laboratory. All data are reported to the detection limit. Results <5x DL must be considered to have a higher degree of uncertainty associated with them. Due to the statistical process of detection limit determination, data in this report should not be used for statistical analysis as the data has been censored in such a manner as to bias statistical analyses high.

---

Project Manager  
Paul Duda

Lab ID: 24-S2462  
Client ID: Run 1 Front Half  
Site: LANL TA-35  
Sample Date: 8/20/24

Analyte	Result	DL	Units
Beryllium, ICP	5.96	0.050	µg/sample

---

Lab ID: 24-S2463  
Client ID: Run 1 Back Half  
Site: LANL TA-35  
Sample Date: 8/20/24

Analyte	Result	DL	Units
Beryllium, ICP	< DL	0.020	µg/sample

---

Lab ID: 24-S2464  
Client ID: Run 2 Front Half  
Site: LANL TA-35  
Sample Date: 8/21/24

Analyte	Result	DL	Units
Beryllium, ICP	5.94	0.050	µg/sample

---

Lab ID: 24-S2465  
Client ID: Run 2 Back Half  
Site: LANL TA-35  
Sample Date: 8/21/24

Analyte	Result	DL	Units
Beryllium, ICP	< DL	0.020	µg/sample

---

Lab ID: 24-S2466  
Client ID: Run 3 Front Half  
Site: LANL TA-35  
Sample Date: 8/21/24

Analyte	Result	DL	Units
Beryllium, ICP	5.94	0.050	µg/sample

---

Lab ID: 24-S2467  
Client ID: Run 3 Back Half  
Site: LANL TA-35  
Sample Date: 8/21/24

Analyte	Result	DL	Units
Beryllium, ICP	< DL	0.020	µg/sample

---

Lab ID: 24-S2468  
Client ID: C12 Filter Blank  
Site: LANL TA-35  
Sample Date: 8/20/24

---

Analyte	Result	DL	Units
Beryllium, ICP	6.01	0.050	µg/sample

---

Lab ID: 24-S2469  
Client ID: C8A HNO3 Blank  
Site: LANL TA-35  
Sample Date: 8/20/24

---

Analyte	Result	DL	Units
Beryllium, ICP	< DL	0.050	µg/sample

---

Lab ID: 24-S2470  
Client ID: C9 5%/10% Blank  
Site: LANL TA-35  
Sample Date: 8/20/24

---

Analyte	Result	DL	Units
Beryllium, ICP	< DL	0.020	µg/sample

---

## QA/QC Report

Client Name: Compliance Services & Testing  
 Project Number: C062  
 Analytical Technique: ICP-OES  
 Instrument: Perkin Elmer Optima 8300  
 Sample Description: Method 29 Front Half  
 Report Number: 24-512

### Blank Data

Analyte	Sample ID	Measured Conc. µg/L	DL Conc. µg/L
Be	ICB	< DL	0.200
Be	Meth_Blk	< DL	0.200
Be	CCB	< DL	0.200
Be	CCB	< DL	0.200

ICB: Initial Calibration Blank CCB: Continuing Calibration Blank  
 \*: Sample Media Blank (SM\_Blk) concentration in µg/filter  
 Method Blank is in control if Method Blank results are <10% of sample results

### Calibration QC

Analyte	Sample ID	Standard Conc. µg/L	Measured Conc. µg/L	Percent Recovery
Be	ICV	2500.	2549.	102.0
Be	LL-CCV	1.000	1.031	103.1
Be	LL-LCS	0.500	0.541	108.2
Be	CCV	2500.	2553.	102.1
Be	CCV	2500.	2534.	101.4

ICV: Initial Calibration Verification CCV: Continuing Calibration Verification  
 Calibration Verification Limits: 90% - 110% Recovery  
 Low Level-CCV Limits: 60% - 140% Recovery  
 Low Level-LCS Limits: 50% - 150% Recovery  
 LL-LCS results are insignificant if sample results are >10x LL-LCS concentration

### Replicate Data

Analyte	Sample ID	Sample Conc. µg/L	Replicate Conc. µg/L	RPD
Be	24-S2462	23.9	24.0	0.67

RPD =  $\frac{\text{sample} - \text{replicate}}{[(\text{sample} + \text{replicate}) / 2]} \times 100$   
 N/C: RPD is not calculated when sample or replicate is below detection limit  
 Replicate Limit: 20% RPD  
 \*: per EPA CLP protocol, control limits do not apply if sample and/or replicate concentration is less than 5x the detection limit

### Laboratory Control Sample/Matrix Post Spike Analysis

Analyte	Sample ID	Sample Conc. µg/L	Spike Conc. µg/L	Spike Amount µg/L	Percent Recovery
Be	LCS	< 0.2	1050.	1000	105.

LCS Limit: 80% - 120% Recovery Spike Limit: 75% - 125% Recovery

## QA/QC Report

Client Name: Compliance Services & Testing  
Project Number: C062  
Analytical Technique: ICP-OES  
Instrument: Perkin Elmer Optima 8300  
Sample Description: Method 29 Front Half  
Report Number: 24-512

### Laboratory Control Sample/Matrix Post Spike Analysis (continued)

Analyte	Sample ID	Sample Conc. µg/L	Spike Conc. µg/L	Spike Amount µg/L	Percent Recovery
Be	LCS-Duplicate	< 0.2	1077.	1000	108.
Be	24-S2464	23.77	932.5	1000	90.9

LCS Limit: 80% - 120% Recovery

Spike Limit: 75% - 125% Recovery

### LCS Duplicate Data

Analyte	Sample ID	Original Conc. µg/L	Replicate Conc. µg/L	RPD
Be	LCS-Duplicate	1050.	1077.	2.54

RPD =  $\frac{(\text{sample} - \text{duplicate})}{[(\text{sample} + \text{duplicate})/2]} \times 100$

Duplicate Limit: 20% RPD

## QA/QC Report

Client Name: Compliance Services & Testing  
 Project Number: C062  
 Analytical Technique: ICP-OES  
 Instrument: Perkin Elmer Optima 8300  
 Sample Description: Method 29 Back Half  
 Report Number: 24-512

### Blank Data

Analyte	Sample ID	Measured Conc. µg/L	DL Conc. µg/L
Be	ICB	< DL	0.200
Be	Meth_Blk	< DL	0.200
Be	CCB	< DL	0.200
Be	CCB	< DL	0.200

ICB: Initial Calibration Blank CCB: Continuing Calibration Blank  
 \*: Sample Media Blank (SM\_Blk) concentration in µg/filter  
 Method Blank is in control if Method Blank results are <10% of sample results

### Calibration QC

Analyte	Sample ID	Standard Conc. µg/L	Measured Conc. µg/L	Percent Recovery
Be	ICV	2500.	2471.	98.8
Be	LL-CCV	1.000	0.979	97.9
Be	LL-LCS	0.500	0.444	88.8
Be	CCV	2500.	2476.	99.0
Be	CCV	2500.	2490.	99.6

ICV: Initial Calibration Verification CCV: Continuing Calibration Verification  
 Calibration Verification Limits: 90% - 110% Recovery  
 Low Level-CCV Limits: 60% - 140% Recovery  
 Low Level-LCS Limits: 50% - 150% Recovery  
 LL-LCS results are insignificant if sample results are >10x LL-LCS concentration

### Replicate Data

Analyte	Sample ID	Sample Conc. µg/L	Replicate Conc. µg/L	RPD
Be	24-S2463	< 0.2	< 0.2	N/C *

RPD =  $\frac{(\text{sample} - \text{replicate})}{[(\text{sample} + \text{replicate})/2]} \times 100$   
 N/C: RPD is not calculated when sample or replicate is below detection limit  
 Replicate Limit: 20% RPD  
 \*: per EPA CLP protocol, control limits do not apply if sample and/or replicate concentration is less than 5x the detection limit

### Laboratory Control Sample/Matrix Post Spike Analysis

Analyte	Sample ID	Sample Conc. µg/L	Spike Conc. µg/L	Spike Amount µg/L	Percent Recovery
Be	LCS	< 0.2	973.9	1000	97.4

LCS Limit: 80% - 120% Recovery Spike Limit: 75% - 125% Recovery



## QA/QC Report

Client Name: Compliance Services & Testing  
Project Number: C062  
Analytical Technique: ICP-OES  
Instrument: Perkin Elmer Optima 8300  
Sample Description: Method 29 Back Half  
Report Number: 24-512

### Laboratory Control Sample/Matrix Post Spike Analysis (continued)

Analyte	Sample ID	Sample Conc. µg/L	Spike Conc. µg/L	Spike Amount µg/L	Percent Recovery
Be	LCS-Duplicate	< 0.2	971.5	1000	97.2
Be	24-S2465	< 0.2	930.3	1000	93.0

LCS Limit: 80% - 120% Recovery

Spike Limit: 75% - 125% Recovery

### LCS Duplicate Data

Analyte	Sample ID	Original Conc. µg/L	Replicate Conc. µg/L	RPD
Be	LCS-Duplicate	973.9	971.5	0.25

RPD =  $\frac{(\text{sample} - \text{duplicate})}{[(\text{sample} + \text{duplicate})/2]} \times 100$

Duplicate Limit: 20% RPD

## Alliance Technical Group SOURCE SAMPLE RECEIPT CHECKLIST

Client Compliance Services & Testing Date 8/23/2024  
 # Runs 3 + blanks Report # 24-512

Package Intact?

Chain-of-Custody form inspected   
 CoC present with samples?   
 CoC indicates analytical methodology to be used? (eg M29, etc.)  M29 !!  
 Has CoC been signed by client?   
 Custody release date and time noted on CoC?

All sample containers inspected   
 Does number of samples match number on CoC form?  !!  
 Do all sample ID numbers match those on the CoC form?  !!  
 Did client mark sample volumes prior to shipment?   
 Sample temperature recorded?  NO Ambient !!  
 Are the sample containers intact?  !!  
 If present, Audit Sample intact?  n/a !!  
 Are signs of leakage present?  None \*

Chain-of-Custody form signed and dated by CLN

Corrective actions   
 Client contacted due to mismatching sample ID numbers  
 Client contacted due to broken sample container(s)  
 Client contacted due to leaking sample container(s)  
 Client contacted for verification of methodology?  
 Corrective actions documented?  
 Corrective actions accomplished?

Items marked *!!* shall be addressed prior to any analytical work being started.  
 Items marked *\** shall be noted in case narrative upon reporting of results to client.

Signed 

Notes \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

### CHAIN-OF-CUSTODY RECORD

For use by Lab:  
Report #: 24-512

Page 1 of 2 COCs

**CLIENT INFORMATION**

Company Name: Compliance Services & Testing (CST)

Contact: CHRIS SPENCER Email: cspencer@comptest.com

Cell: 503-681-4909 Office: SAME

Report To: SAME email Billing Address: SAME

**Analysis Requested**

Gravimetry	XRF	IC	ICP	OC/EC	Cr VI
			Method 29		

Turn Around Time Requested

Standard  
 Rush (Specify)

LOWEST  
DETECTION  
LIMITS

Project Name: LANL TA-35 Quote Number: \_\_\_\_\_  
PO# 2634

Lab ID	Field Sample ID	Site	Sample Date	Volume (m <sup>3</sup> )	Particle Size
QF-16	FILTER BLANK	# 2 unit	8-20-24	N/A	N/A
QF-15	RUN 1		8-20-24		
QF-17	RUN 2		8-24-24		
QF-18	RUN 3		8-21-24		
A	0.1% HNO <sub>3</sub> / 10% H <sub>2</sub> O <sub>2</sub> BLANK		8-20-24	100 mL	
B	0.1% HNO <sub>3</sub> BLANK		8-20-24	100 mL	
IA	0.1% NITRIC RINSE	RUN #1		TBD	
IB	0.1% NITRIC RINSE				
IC	IMMERSE CATCH				

Sample Specific Notes

0.1% HNO<sub>3</sub> / 10% H<sub>2</sub>O<sub>2</sub> / 85% DI

0.1% HNO<sub>3</sub> / H<sub>2</sub>O<sub>2</sub> IMM CATCH  
+ SAME 0.1% HNO<sub>3</sub> RINSE

Do the samples pose any potential hazards?  
If yes please explain:  Yes  No

Are samples for compliance?  Yes  No

Special Instructions/QC Requirements & Comments:

Relinquished by: CHRIS SPENCER

Date/Time: 8-22-2024

Received By: [Signature]

Date/Time/Temp: 8/23/24 9:05

Laboratory Receipt Comments:

## CHAIN-OF-CUSTODY RECORD

For use by Lab:  
 Report #: 24-512

Page 2 of 2 COCs

**CLIENT INFORMATION**

Company Name: CST

Contact: CHRIS SPENCER Email: cspencer@comp-testing.com

Cell: 503-681-4909 Office: SAME

Report To: Same Billing Address: SAME

**Analysis Requested**

Turn Around Time Requested

Standard  
 Rush (Specify)

Project Name: LANL TA-35 Quote Number:

PO#: 2634

Lab ID	Field Sample ID	Site	Sample Date	Volume (m <sup>3</sup> )	Particle Size	Gravimetry	XRF	IC	ICP METHOD 29	OC/EC	Cr VI
2A	NITRIC RINSE	RUN 2	5-21-24	TBD	N/A	24-S2464					
2B	NITRIC RINSE	↓	↓	↓	↓	24-S2465					
2C	IMPINGER	↓	↓	↓	↓						
3A	NITRIC RINSE	RUN 3	↓	↓	↓	24-S2466					
3B	NITRIC RINSE	↓	↓	↓	↓	24-S2467					
3C	IMPINGER	↓	↓	↓	↓						

Sample Specific Notes:

0.1N HNO<sub>3</sub>  
 (6.3 mL of 70% HNO<sub>3</sub> → 1 L DI)  
 5% HNO<sub>3</sub> / 10% H<sub>2</sub>O<sub>2</sub>  
 0.1N HNO<sub>3</sub>  
 0.1N HNO<sub>3</sub>  
 5% HNO<sub>3</sub> / 10% H<sub>2</sub>O<sub>2</sub> / 85% DI

Do the samples pose any potential hazards?  Yes  No

Are samples for compliance?  Yes  No

Special Instructions/QC Requirements & Comments:

Relinquished by: [Signature] Date/Time: 08-22-2024 Received By: [Signature] Date/Time/Temp: 8/23/24 9.05

Laboratory Receipt Comments:

Method 29 Digestion Worklist Number: 2834

Label QC by: 8-26-24 SG

Report #: 24-512

Client: C062

Compliance Services & Testing

Digestion Date:

<u>Bomb</u>	Sample ID	Client ID	Digestate Volume (ml)	pH	Notes
15	1. MethBlk		<u>250</u>		
18	2. LL-LCS19				2.5 ml 248.63.5 X:11.11.24
19	3. LCS19				2.5 ml 100ppm LCS19
21	4. LCS_Dup19				2.5 ml 100ppm LCS19
22	5. 24-S2462	Run 1 C1&3			
23	6. 24-S2464	Run 2 C1&3			
24	7. 24-S2466	Run 3 C1&3			
25	8. 24-S2468	C12 Filter Blan			100 ml used
26	9. 24-S2469	C8A HNO3 Blank			100 ml used

Be

•4ml HF per bomb

Digestion Method: M29\_FH M29\_BH CARB436\_Combined Other: \_\_\_\_\_

Volumetrics: HNO3 10-2, 5.0 ml, 5.0 G, HF250-1, HF250-2

Evap Start: 8-26-24 SG

HNO3 Lot: omni trace 63320 X:11/6

Evap End: 8-27-24 SG

H2O2 Lot: \_\_\_\_\_

Digest: 8-28-24 KH

HF Lot: CP V093001 X:9-4-25

Dilute: 8-28-24 KH

Spike Std: Accu std 223055011  
X:6-4-25

ICP  
Client: C062  
Report #: 24-512

Worklist Number: 2944  
Compliance Services & Testing

Date Requested: 8/29/24  
Analysis Date: 8/29/24

Q COK  
8/30/24

Lab ID	Analyte Results in ug/L	Expected	% Recovery
1. kv:ICV	Be 2549. ✓	2500.	101.96
2. bl:ICB	Be < 0.2 ✓		
3. kv:LL-CCV	Be 1.031 ✓	Expected 1.000	% Recovery 103.10
4. bl:Meth_Blk	Be < 0.2 ✓		
5. kv:LL-LCS	Be 0.541 ✓	Expected 0.500	% Recovery 108.20
6. Spike of 4 ✓	Be 1050. ✓	Spk Amount 1000.	Spike Rec. 105.00
7. MSD of 6 ✓	Be 1077. ✓	Recovery 107.70	RPD 2.54
8. 24-S2462 Run 1 C1&3	Be 23.86 ✓		
9. Dup of 8	Be 24.02 ✓	Difference 0.160	RPD 0.67
10. 24-S2464 Run 2 C1&3	Be 23.77 ✓		
11. Spike of 10	Be 932.5 ✓	Spk Amount 1000.	Spike Rec. 90.87
12. 24-S2466 Run 3 C1&3	Be 23.77 ✓		
13. kv:CCV	Be 2553. ✓	Expected 2500.	% Recovery 102.12
14. bl:CCB	Be < 0.2 ✓		
15. 24-S2468 C12 Filter Bl	Be 24.05 ✓		
16. 24-S2469 C8A HNO3 Blan	Be < 0.2 ✓		
17. kv:CCV	Be 2534. ✓	Expected 2500.	% Recovery 101.36
18. bl:CCB	Be < 0.2 ✓		



-----  
Nebulizer Parameters: ICV

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

-----  
Replicate Data: ICV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	11553278.2	11552328.5	2.583 mg/L	2583 ug/L	2:12:30 PM
2	Be 313.107	11409894.8	11408945.2	2.550 mg/L	2550 ug/L	2:12:46 PM
3	Be 313.107	11241825.9	11240876.2	2.513 mg/L	2513 ug/L	2:13:03 PM

-----  
Mean Data: ICV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	11400716.6	2.549 mg/L	0.0348	2549 ug/L	34.85	1.37%

QC value within limits for Be 313.107 Recovery = 101.95%

All analyte(s) passed QC.

## =====

Sequence No.: 4  
 Sample ID: ICB  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 4  
 Date Collected: 8/29/2024 2:15:54 PM  
 Data Type: Reprocessed on 8/29/2024 6:54:31 PM

Initial Sample Vol:

Sample Prep Vol:

-----  
Nebulizer Parameters: ICB

Analyte	Back Pressure	Flow
All	342.0 kPa	0.65 L/min

-----  
Replicate Data: ICB

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1224.6	274.9	0.000 mg/L	0.061 ug/L	2:16:44 PM
2	Be 313.107	1393.2	443.5	0.000 mg/L	0.099 ug/L	2:16:56 PM
3	Be 313.107	1296.2	346.5	0.000 mg/L	0.077 ug/L	2:17:08 PM

-----  
Mean Data: ICB

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	355.0	0.000 mg/L	0.0000	0.079 ug/L	0.0189	23.85%

QC value within limits for Be 313.107 Recovery = Not calculated

All analyte(s) passed QC.

## =====

Sequence No.: 5  
 Sample ID: LL-CCV  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 5  
 Date Collected: 8/29/2024 2:19:57 PM  
 Data Type: Reprocessed on 8/29/2024 6:54:31 PM

Initial Sample Vol:

Sample Prep Vol:

-----  
Nebulizer Parameters: LL-CCV

Analyte	Back Pressure	Flow
All	342.0 kPa	0.65 L/min

-----  
Replicate Data: LL-CCV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	5751.5	4801.8	0.001 mg/L	1.073 ug/L	2:20:48 PM
2	Be 313.107	5507.2	4557.5	0.001 mg/L	1.019 ug/L	2:21:00 PM
3	Be 313.107	5422.4	4472.7	0.001 mg/L	1.000 ug/L	2:21:12 PM



-----  
Mean Data: LL-CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4610.7	0.001 mg/L	0.0000	1.031 ug/L	0.0382	3.71%

QC value within limits for Be 313.107 Recovery = 103.07%  
All analyte(s) passed QC.

Sequence No.: 6  
Sample ID: MB 24-512 FH  
Analyst: SG  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt:  
Dilution:  
Wash Time:

Autosampler Location: 14  
Date Collected: 8/29/2024 2:24:01 PM  
Data Type: Reprocessed on 8/29/2024 6:54:31 PM  
Initial Sample Vol:  
Sample Prep Vol:

-----  
Nebulizer Parameters: MB 24-512 FH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

-----  
Replicate Data: MB 24-512 FH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1047.3	97.6	0.000 mg/L	0.022 ug/L	2:24:51 PM
2	Be 313.107	1322.2	372.5	0.000 mg/L	0.083 ug/L	2:25:04 PM
3	Be 313.107	1379.6	429.9	0.000 mg/L	0.096 ug/L	2:25:16 PM

-----  
Mean Data: MB 24-512 FH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	300.0	0.000 mg/L	0.0000	0.067 ug/L	0.0397	59.21%

Sequence No.: 7  
Sample ID: LL-LCS 24-512 FH  
Analyst: SG  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt:  
Dilution:  
Wash Time:

Autosampler Location: 15  
Date Collected: 8/29/2024 2:28:04 PM  
Data Type: Reprocessed on 8/29/2024 6:54:31 PM  
Initial Sample Vol:  
Sample Prep Vol:

-----  
Nebulizer Parameters: LL-LCS 24-512 FH

Analyte	Back Pressure	Flow
All	342.0 kPa	0.65 L/min

-----  
Replicate Data: LL-LCS 24-512 FH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	3379.7	2430.0	0.001 mg/L	0.543 ug/L	2:28:53 PM
2	Be 313.107	3364.3	2414.6	0.001 mg/L	0.540 ug/L	2:29:06 PM
3	Be 313.107	3369.6	2419.9	0.001 mg/L	0.541 ug/L	2:29:18 PM

-----  
Mean Data: LL-LCS 24-512 FH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	2421.5	0.001 mg/L	0.0000	0.541 ug/L	0.0018	0.32%

Sequence No.: 8  
Sample ID: LCS 24-512 FH  
Analyst: SG  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt:  
Dilution:

Autosampler Location: 16  
Date Collected: 8/29/2024 2:32:05 PM  
Data Type: Reprocessed on 8/29/2024 6:54:32 PM  
Initial Sample Vol:  
Sample Prep Vol:

Wash Time:

Nebulizer Parameters: LCS 24-512 FH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

Replicate Data: LCS 24-512 FH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4714259.1	4713309.4	1.054 mg/L	1054 ug/L	2:32:57 PM
2	Be 313.107	4697560.3	4696610.6	1.050 mg/L	1050 ug/L	2:33:12 PM
3	Be 313.107	4683816.3	4682866.6	1.047 mg/L	1047 ug/L	2:33:27 PM

Mean Data: LCS 24-512 FH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4697595.5	1.050 mg/L	0.0034	1050 ug/L	3.41	0.32%

Sequence No.: 9

Sample ID: LCS DUP 24-512 FH

Analyst: SG

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Dilution:

Wash Time:

Autosampler Location: 17

Date Collected: 8/29/2024 2:36:16 PM

Data Type: Reprocessed on 8/29/2024 6:54:32 PM

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: LCS DUP 24-512 FH

Analyte	Back Pressure	Flow
All	342.0 kPa	0.65 L/min

Replicate Data: LCS DUP 24-512 FH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4858463.9	4857514.2	1.086 mg/L	1086 ug/L	2:37:06 PM
2	Be 313.107	4801271.3	4800321.6	1.073 mg/L	1073 ug/L	2:37:23 PM
3	Be 313.107	4797431.2	4796481.5	1.072 mg/L	1072 ug/L	2:37:41 PM

Mean Data: LCS DUP 24-512 FH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4818105.8	1.077 mg/L	0.0076	1077 ug/L	7.64	0.71%

Sequence No.: 10

Sample ID: 24-S2462

Analyst: SG

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Dilution:

Wash Time:

Autosampler Location: 18

Date Collected: 8/29/2024 2:40:32 PM

Data Type: Reprocessed on 8/29/2024 6:54:32 PM

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: 24-S2462

Analyte	Back Pressure	Flow
All	342.0 kPa	0.65 L/min

Replicate Data: 24-S2462

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	108103.8	107154.1	0.024 mg/L	23.95 ug/L	2:41:29 PM
2	Be 313.107	107634.3	106684.6	0.024 mg/L	23.85 ug/L	2:41:47 PM
3	Be 313.107	107283.1	106333.4	0.024 mg/L	23.77 ug/L	2:42:05 PM

Mean Data: 24-S2462

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	106724.0	0.024 mg/L		0.0001	23.86 ug/L	0.092	0.39%

Sequence No.: 11  
 Sample ID: 24-S2462 REP  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 19  
 Date Collected: 8/29/2024 2:44:52 PM  
 Data Type: Reprocessed on 8/29/2024 6:54:32 PM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2462 REP

Analyte	Back Pressure	Flow
All	341.0 kPa	0.65 L/min

Replicate Data: 24-S2462 REP

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	110197.5	109247.8	0.024 mg/L	24.42 ug/L	2:45:49 PM
2	Be 313.107	106958.2	106008.5	0.024 mg/L	23.70 ug/L	2:46:08 PM
3	Be 313.107	108071.2	107121.5	0.024 mg/L	23.95 ug/L	2:46:26 PM

Mean Data: 24-S2462 REP

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	107459.3	0.024 mg/L		0.0004	24.02 ug/L	0.368	1.53%

Sequence No.: 12  
 Sample ID: 24-S2464  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 20  
 Date Collected: 8/29/2024 2:49:14 PM  
 Data Type: Reprocessed on 8/29/2024 6:54:32 PM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2464

Analyte	Back Pressure	Flow
All	341.0 kPa	0.65 L/min

Replicate Data: 24-S2464

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	108637.1	107687.4	0.024 mg/L	24.07 ug/L	2:50:10 PM
2	Be 313.107	106181.8	105232.1	0.024 mg/L	23.52 ug/L	2:50:28 PM
3	Be 313.107	106961.0	106011.3	0.024 mg/L	23.70 ug/L	2:50:46 PM

Mean Data: 24-S2464

Analyte	Mean Corrected Intensity	Conc. Units	Calib.	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	106310.3	0.024 mg/L		0.0003	23.77 ug/L	0.280	1.18%

Sequence No.: 13  
 Sample ID: 24-S2464 POST SPK  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 21  
 Date Collected: 8/29/2024 2:53:35 PM  
 Data Type: Reprocessed on 8/29/2024 6:54:32 PM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2464 POST SPK

Analyte Back Pressure Flow  
 All 342.0 kPa 0.65 L/min

-----  
 Replicate Data: 24-S2464 POST SPK

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4211283.8	4210334.1	0.941 mg/L	941.2 ug/L	2:54:30 PM
2	Be 313.107	4129196.6	4128246.9	0.923 mg/L	922.9 ug/L	2:54:49 PM
3	Be 313.107	4176321.6	4175371.9	0.933 mg/L	933.4 ug/L	2:55:07 PM

-----  
 Mean Data: 24-S2464 POST SPK

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4171317.6	0.932 mg/L	0.0092	932.5 ug/L	9.21	0.99%

=====

Sequence No.: 14	Autosampler Location: 22
Sample ID: 24-S2466	Date Collected: 8/29/2024 2:57:57 PM
Analyst: SG	Data Type: Reprocessed on 8/29/2024 6:54:32 PM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
 Nebulizer Parameters: 24-S2466

Analyte Back Pressure Flow  
 All 342.0 kPa 0.65 L/min

-----  
 Replicate Data: 24-S2466

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	108472.7	107523.0	0.024 mg/L	24.04 ug/L	2:58:56 PM
2	Be 313.107	107389.5	106439.8	0.024 mg/L	23.79 ug/L	2:59:14 PM
3	Be 313.107	106041.7	105092.0	0.023 mg/L	23.49 ug/L	2:59:32 PM

-----  
 Mean Data: 24-S2466

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	106351.6	0.024 mg/L	0.0003	23.77 ug/L	0.272	1.15%

=====

Sequence No.: 15	Autosampler Location: 3
Sample ID: CCV	Date Collected: 8/29/2024 3:02:21 PM
Analyst: SG	Data Type: Reprocessed on 8/29/2024 6:54:32 PM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
 Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
 All 342.0 kPa 0.65 L/min

-----  
 Replicate Data: CCV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	11427928.3	11426978.6	2.554 mg/L	2554 ug/L	3:03:14 PM
2	Be 313.107	11455289.4	11454339.8	2.561 mg/L	2561 ug/L	3:03:31 PM
3	Be 313.107	11375544.1	11374594.4	2.543 mg/L	2543 ug/L	3:03:47 PM

-----  
 Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	11418637.6	2.553 mg/L	0.0091	2553 ug/L	9.06	0.35%

QC value within limits for Be 313.107 Recovery = 102.11%  
All analyte(s) passed QC.

```
=====
Sequence No.: 16                               Autosampler Location: 4
Sample ID: CCB                                 Date Collected: 8/29/2024 3:06:38 PM
Analyst: SG                                   Data Type: Reprocessed on 8/29/2024 6:54:33 PM
Logged In Analyst (Original) : Administrator
Initial Sample Wt:                             Initial Sample Vol:
Dilution:                                    Sample Prep Vol:
Wash Time:
```

```
-----
Nebulizer Parameters: CCB
Analyte          Back Pressure      Flow
All              342.0 kPa           0.65 L/min
```

## Replicate Data: CCB

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1198.5	248.8	0.000 mg/L	0.056 ug/L	3:07:29 PM
2	Be 313.107	1201.0	251.3	0.000 mg/L	0.056 ug/L	3:07:41 PM
3	Be 313.107	1123.5	173.8	0.000 mg/L	0.039 ug/L	3:07:53 PM

## Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	224.6	0.000 mg/L	0.0000	0.050 ug/L	0.0098	19.61%

QC value within limits for Be 313.107 Recovery = Not calculated  
All analyte(s) passed QC.

```
=====
Sequence No.: 17                               Autosampler Location: 23
Sample ID: 24-S2468                           Date Collected: 8/29/2024 3:10:41 PM
Analyst: SG                                   Data Type: Reprocessed on 8/29/2024 6:54:33 PM
Logged In Analyst (Original) : Administrator
Initial Sample Wt:                             Initial Sample Vol:
Dilution:                                    Sample Prep Vol:
Wash Time:
```

```
-----
Nebulizer Parameters: 24-S2468
Analyte          Back Pressure      Flow
All              342.0 kPa           0.65 L/min
```

## Replicate Data: 24-S2468

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	107574.7	106625.0	0.024 mg/L	23.84 ug/L	3:11:39 PM
2	Be 313.107	109155.3	108205.6	0.024 mg/L	24.19 ug/L	3:11:57 PM
3	Be 313.107	108855.2	107905.6	0.024 mg/L	24.12 ug/L	3:12:15 PM

## Mean Data: 24-S2468

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	107578.7	0.024 mg/L	0.0002	24.05 ug/L	0.188	0.78%

```
=====
Sequence No.: 18                               Autosampler Location: 24
Sample ID: 24-S2469                           Date Collected: 8/29/2024 3:15:03 PM
Analyst: SG                                   Data Type: Reprocessed on 8/29/2024 6:54:33 PM
Logged In Analyst (Original) : Administrator
Initial Sample Wt:                             Initial Sample Vol:
Dilution:                                    Sample Prep Vol:
Wash Time:
```

```
-----
Nebulizer Parameters: 24-S2469
```

Analyte Back Pressure Flow  
 All 342.0 kPa 0.65 L/min

-----  
 Replicate Data: 24-S2469

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	959.7	10.0	0.000 mg/L	0.002 ug/L	3:15:55 PM
2	Be 313.107	994.8	45.1	0.000 mg/L	0.010 ug/L	3:16:07 PM
3	Be 313.107	914.0	-35.7	-0.000 mg/L	-0.008 ug/L	3:16:19 PM

-----  
 Mean Data: 24-S2469

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	6.5	0.000 mg/L	0.0000	0.001 ug/L	0.0091	626.44%

=====

Sequence No.: 19	Autosampler Location: 3
Sample ID: CCV	Date Collected: 8/29/2024 3:19:08 PM
Analyst: SG	Data Type: Reprocessed on 8/29/2024 6:54:33 PM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
 Nebulizer Parameters: CCV

Analyte Back Pressure Flow  
 All 343.0 kPa 0.65 L/min

-----  
 Replicate Data: CCV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	11403153.5	11402203.8	2.549 mg/L	2549 ug/L	3:19:59 PM
2	Be 313.107	11322732.7	11321783.0	2.531 mg/L	2531 ug/L	3:20:16 PM
3	Be 313.107	11288045.6	11287095.9	2.523 mg/L	2523 ug/L	3:20:32 PM

-----  
 Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	11337027.6	2.534 mg/L	0.0132	2534 ug/L	13.20	0.52%

QC value within limits for Be 313.107 Recovery = 101.38%  
 All analyte(s) passed QC.

=====

Sequence No.: 20	Autosampler Location: 4
Sample ID: CCB	Date Collected: 8/29/2024 3:23:23 PM
Analyst: SG	Data Type: Reprocessed on 8/29/2024 6:54:33 PM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
 Nebulizer Parameters: CCB

Analyte Back Pressure Flow  
 All 342.0 kPa 0.65 L/min

-----  
 Replicate Data: CCB

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1198.9	249.2	0.000 mg/L	0.056 ug/L	3:24:15 PM
2	Be 313.107	1102.6	152.9	0.000 mg/L	0.034 ug/L	3:24:27 PM
3	Be 313.107	1121.1	171.4	0.000 mg/L	0.038 ug/L	3:24:39 PM

-----  
 Mean Data: CCB

Mean Corrected Intensity	Calib. Conc. Units	Sample Conc. Units
--------------------------	--------------------	--------------------

Analyte	Intensity	Conc. Units	Std.Dev.	Conc. Units	Std.Dev.	RSD
Be 313.107	191.2	0.000 mg/L	0.0000	0.043 ug/L	0.0114	26.73%

QC value within limits for Be 313.107 Recovery = Not calculated

All analyte(s) passed QC.

Method 29 Digestion Worklist Number: 2835  
 Report #: 24-512 Client: C062

Label QC by: B.27-2486  
 Compliance Services & Testing  
 Digestion Date:

	Sample ID	Client ID	Digestate Volume (ml)	pH	Notes
1.	MethBlk		<u>100</u>		
2.	LL-LCS19		}		1.0mL 248.635 X:11.11.24
3.	LCS19				1.0mL 100 ppm LCAP 19
4.	LCS_Dup19				1.0mL 100 ppm LCAP 19
5.	24-S2463	Run 1 BH&C4			
6.	24-S2465	Run 2 BH&C4			
7.	24-S2467	Run 3 BH&C4			
8.	24-S2470	C9 5%/10% Blank	↓		

Bc

Digestion Method: M29\_FH M29\_BH CARB436\_Combined Other: \_\_\_\_\_

Volumetrics: HNO3 10-2, 1.0i, M100-1

Evap Start: 8.27.24 SG

HNO3 Lot: omni trace 63320 X:NG

Evap End: 9.4.24 SG

H2O2 Lot: CP V029001 X:12.11.24

Digest: 9.4.24 SG

HF Lot: \_\_\_\_\_

Dilute: 9.4.24 SG

Spike Std: AccuStd. 223055011

X:16.4.25



*QC011  
9.5.24*

Lab ID	Analyte	Results in ug/L	Expected	% Recovery
1. kv:ICV	Be	2471. ✓	2500.	98.84
2. bl:ICB	Be	< 0.2 ✓		
3. kv:LL-CCV	Be	0.979 ✓	Expected 1.000	% Recovery 97.90
4. bl:Meth_Bl	Be	< 0.2 ✓		
5. kv:LL-LCS	Be	0.444 ✓	Expected 0.500	% Recovery 88.80
6. Spike of 4	Be	973.9 ✓	Spk Amount 1000.	Spike Rec. 97.39
7. MSD of 6	Be	971.5 ✓	Recovery 97.15	RPD -0.25
8. 24-S2463 Run 1 BH&C4	Be	< 0.2 ✓		
9. Dup of 8	Be	< 0.2 ✓	Difference	RPD
10. 24-S2465 Run 2 BH&C4	Be	< 0.2 ✓		
11. Spike of 10	Be	930.3 ✓	Spk Amount 1000.	Spike Rec. 93.03
12. 24-S2467 Run 3 BH&C4	Be	< 0.2 ✓		
13. kv:CCV	Be	2476. ✓	Expected 2500.	% Recovery 99.04
14. bl:CCB	Be	< 0.2 ✓		
15. 24-S2470 C9 5%/10% Bla	Be	< 0.2 ✓		
16. kv:CCV	Be	2490. ✓	Expected 2500.	% Recovery 99.60
17. bl:CCB	Be	< 0.2 ✓		

```

Sequence No.: 1
Sample ID: Calib Blank 1
Analyst:
Logged In Analyst (Original) : Administrator
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 1
Date Collected: 9/5/2024 6:48:34 AM
Data Type: Reprocessed on 9/5/2024 8:32:14 AM

Initial Sample Vol:
Sample Prep Vol:
    
```

```

Nebulizer Parameters: Calib Blank 1
Analyte      Back Pressure  Flow
All          339.0 kPa     0.65 L/min
    
```

```

Replicate Data: Calib Blank 1
Repl# Analyte      Net      Corrected      Calib.      Analysis
      Analyte      Intensity  Intensity      Conc. Units  Time
1 Be 313.107      1131.3     1131.3         [0.00] mg/L  6:49:23 AM
2 Be 313.107      1204.9     1204.9         [0.00] mg/L  6:49:35 AM
3 Be 313.107      1163.7     1163.7         [0.00] mg/L  6:49:47 AM
    
```

```

Mean Data: Calib Blank 1
Analyte      Mean Corrected      Std.Dev.  RSD      Calib
      Analyte      Intensity          Std.Dev.  RSD      Conc. Units
Be 313.107      1166.6           36.92     3.16%     [0.00] mg/L
    
```

```

Sequence No.: 2
Sample ID: Calib Std 1
Analyst:
Logged In Analyst (Original) : Administrator
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 2
Date Collected: 9/5/2024 6:52:35 AM
Data Type: Reprocessed on 9/5/2024 8:32:14 AM

Initial Sample Vol:
Sample Prep Vol:
    
```

```

Nebulizer Parameters: Calib Std 1
Analyte      Back Pressure  Flow
All          341.0 kPa     0.65 L/min
    
```

```

Replicate Data: Calib Std 1
Repl# Analyte      Net      Corrected      Calib.      Analysis
      Analyte      Intensity  Intensity      Conc. Units  Time
1 Be 313.107      23614420.9  23613254.3     [5.0000] mg/L  6:53:28 AM
2 Be 313.107      23524261.5  23523094.9     [5.0000] mg/L  6:53:39 AM
3 Be 313.107      23494621.7  23493455.1     [5.0000] mg/L  6:53:49 AM
    
```

```

Mean Data: Calib Std 1
Analyte      Mean Corrected      Std.Dev.  RSD      Calib
      Analyte      Intensity          Std.Dev.  RSD      Conc. Units
Be 313.107      23543268.1       62395.36  0.27%     [5.0000] mg/L
    
```

Calibration Summary

Analyte	Stds.	Equation	Intercept	Slope	Curvature	Corr. Coef.	Reslope
Be 313.107	1	Lin, Calc Int	0.0	4709000	0.00000	1.000000	

```

Sequence No.: 3
Sample ID: ICV
Analyst: SG
Logged In Analyst (Original) : Administrator
Initial Sample Wt:
Dilution:
Wash Time:

Autosampler Location: 3
Date Collected: 9/5/2024 6:56:33 AM
Data Type: Reprocessed on 9/5/2024 8:32:14 AM

Initial Sample Vol:
Sample Prep Vol:
    
```

Nebulizer Parameters: ICV

Analyte Back Pressure Flow
All 341.0 kPa 0.65 L/min

Replicate Data: ICV

Table with 7 columns: Repl#, Analyte, Net Intensity, Corrected Intensity, Calib. Conc. Units, Sample Conc. Units, Analysis Time. Contains 3 rows of replicate data for Be 313.107.

Mean Data: ICV

Analyte Mean Corrected Intensity Calib. Conc. Units Std.Dev. Sample Conc. Units Std.Dev. RSD
Be 313.107 11637367.5 2.471 mg/L 0.0168 2471 ug/L 16.84 0.68%

QC value within limits for Be 313.107 Recovery = 98.86%
All analyte(s) passed QC.

Sequence No.: 4

Sample ID: ICB

Analyst: SG

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Dilution:

Wash Time:

Autosampler Location: 4

Date Collected: 9/5/2024 7:00:49 AM

Data Type: Reprocessed on 9/5/2024 8:32:14 AM

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: ICB

Analyte Back Pressure Flow
All 341.0 kPa 0.65 L/min

Replicate Data: ICB

Table with 7 columns: Repl#, Analyte, Net Intensity, Corrected Intensity, Calib. Conc. Units, Sample Conc. Units, Analysis Time. Contains 3 rows of replicate data for Be 313.107.

Mean Data: ICB

Analyte Mean Corrected Intensity Calib. Conc. Units Std.Dev. Sample Conc. Units Std.Dev. RSD
Be 313.107 311.6 0.000 mg/L 0.0000 0.066 ug/L 0.0290 43.79%

QC value within limits for Be 313.107 Recovery = Not calculated
All analyte(s) passed QC.

Sequence No.: 5

Sample ID: LL-CCV

Analyst: SG

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Dilution:

Wash Time:

Autosampler Location: 5

Date Collected: 9/5/2024 7:04:53 AM

Data Type: Reprocessed on 9/5/2024 8:32:14 AM

Initial Sample Vol:

Sample Prep Vol:

Nebulizer Parameters: LL-CCV

Analyte Back Pressure Flow
All 342.0 kPa 0.65 L/min

Replicate Data: LL-CCV

Table with 7 columns: Repl#, Analyte, Net Intensity, Corrected Intensity, Calib. Conc. Units, Sample Conc. Units, Analysis Time. Contains 3 rows of replicate data for Be 313.107.

-----  
Mean Data: LL-CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4610.0	0.001 mg/L	0.0000	0.979 ug/L	0.0028	0.29%

QC value within limits for Be 313.107 Recovery = 97.90%  
All analyte(s) passed QC.

Sequence No. : 6	Autosampler Location: 14
Sample ID: MB 24-512 BH	Date Collected: 9/5/2024 7:08:58 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:15 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
Nebulizer Parameters: MB 24-512 BH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

-----  
Replicate Data: MB 24-512 BH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1123.7	-43.0	-0.000 mg/L	-0.009 ug/L	7:09:47 AM
2	Be 313.107	1006.9	-159.8	-0.000 mg/L	-0.034 ug/L	7:09:59 AM
3	Be 313.107	974.1	-192.5	-0.000 mg/L	-0.041 ug/L	7:10:11 AM

-----  
Mean Data: MB 24-512 BH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	-131.8	-0.000 mg/L	0.0000	-0.028 ug/L	0.0167	59.67%

Sequence No. : 7	Autosampler Location: 15
Sample ID: LL-LCS 24-512 BH	Date Collected: 9/5/2024 7:12:59 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:15 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
Nebulizer Parameters: LL-LCS 24-512 BH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

-----  
Replicate Data: LL-LCS 24-512 BH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	3187.0	2020.4	0.000 mg/L	0.429 ug/L	7:13:48 AM
2	Be 313.107	3229.3	2062.6	0.000 mg/L	0.438 ug/L	7:14:00 AM
3	Be 313.107	3349.2	2182.6	0.000 mg/L	0.464 ug/L	7:14:12 AM

-----  
Mean Data: LL-LCS 24-512 BH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	2088.5	0.000 mg/L	0.0000	0.444 ug/L	0.0179	4.03%

Sequence No. : 8	Autosampler Location: 16
Sample ID: LCS 24-512 BH	Date Collected: 9/5/2024 7:17:00 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:15 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:

Wash Time:

Nebulizer Parameters: LCS 24-512 BH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

Replicate Data: LCS 24-512 BH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4649256.8	4648090.2	0.987 mg/L	987.1 ug/L	7:17:50 AM
2	Be 313.107	4555790.2	4554623.6	0.967 mg/L	967.3 ug/L	7:18:05 AM
3	Be 313.107	4556255.2	4555088.6	0.967 mg/L	967.4 ug/L	7:18:20 AM

Mean Data: LCS 24-512 BH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4585934.1	0.974 mg/L	0.0114	973.9 ug/L	11.43	1.17%

Sequence No.: 9

Autosampler Location: 17

Sample ID: LCS DUP 24-512 BH

Date Collected: 9/5/2024 7:21:10 AM

Analyst: SG

Data Type: Reprocessed on 9/5/2024 8:32:15 AM

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Wash Time:

Nebulizer Parameters: LCS DUP 24-512 BH

Analyte	Back Pressure	Flow
All	343.0 kPa	0.65 L/min

Replicate Data: LCS DUP 24-512 BH

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4601461.5	4600294.8	0.977 mg/L	977.0 ug/L	7:22:00 AM
2	Be 313.107	4541972.3	4540805.7	0.964 mg/L	964.4 ug/L	7:22:15 AM
3	Be 313.107	4582963.0	4581796.4	0.973 mg/L	973.1 ug/L	7:22:30 AM

Mean Data: LCS DUP 24-512 BH

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4574299.0	0.971 mg/L	0.0065	971.5 ug/L	6.47	0.67%

Sequence No.: 10

Autosampler Location: 18

Sample ID: 24-S2463

Date Collected: 9/5/2024 7:25:20 AM

Analyst: SG

Data Type: Reprocessed on 9/5/2024 8:32:15 AM

Logged In Analyst (Original) : Administrator

Initial Sample Wt:

Initial Sample Vol:

Dilution:

Sample Prep Vol:

Wash Time:

Nebulizer Parameters: 24-S2463

Analyte	Back Pressure	Flow
All	344.0 kPa	0.65 L/min

Replicate Data: 24-S2463

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1095.5	-71.1	-0.000 mg/L	-0.015 ug/L	7:26:14 AM
2	Be 313.107	977.4	-189.3	-0.000 mg/L	-0.040 ug/L	7:26:31 AM
3	Be 313.107	1020.7	-146.0	-0.000 mg/L	-0.031 ug/L	7:26:48 AM

Mean Data: 24-S2463

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	-135.4	-0.000	mg/L	0.0000	-0.029	ug/L	0.0127 44.12%

Sequence No.: 11  
 Sample ID: 24-S2463 REP  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 19  
 Date Collected: 9/5/2024 7:29:36 AM  
 Data Type: Reprocessed on 9/5/2024 8:32:15 AM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2463 REP

Analyte	Back Pressure	Flow
All	344.0 kPa	0.65 L/min

Replicate Data: 24-S2463 REP

Repl#	Analyte	Net Intensity	Corrected Intensity	Conc. Units	Calib. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	930.8	-235.8	-0.000	mg/L	-0.050	ug/L 7:30:30 AM
2	Be 313.107	985.2	-181.5	-0.000	mg/L	-0.039	ug/L 7:30:47 AM
3	Be 313.107	1086.3	-80.4	-0.000	mg/L	-0.017	ug/L 7:31:03 AM

Mean Data: 24-S2463 REP

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	-165.9	-0.000	mg/L	0.0000	-0.035	ug/L	0.0168 47.55%

Sequence No.: 12  
 Sample ID: 24-S2465  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 20  
 Date Collected: 9/5/2024 7:33:51 AM  
 Data Type: Reprocessed on 9/5/2024 8:32:15 AM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2465

Analyte	Back Pressure	Flow
All	344.0 kPa	0.65 L/min

Replicate Data: 24-S2465

Repl#	Analyte	Net Intensity	Corrected Intensity	Conc. Units	Calib. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	991.4	-175.3	-0.000	mg/L	-0.037	ug/L 7:34:44 AM
2	Be 313.107	888.7	-277.9	-0.000	mg/L	-0.059	ug/L 7:34:58 AM
3	Be 313.107	908.2	-258.4	-0.000	mg/L	-0.055	ug/L 7:35:13 AM

Mean Data: 24-S2465

Analyte	Mean Corrected Intensity	Conc. Units	Calib. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	-237.2	-0.000	mg/L	0.0000	-0.050	ug/L	0.0116 22.98%

Sequence No.: 13  
 Sample ID: 24-S2465 POST SPK  
 Analyst: SG  
 Logged In Analyst (Original) : Administrator  
 Initial Sample Wt:  
 Dilution:  
 Wash Time:

Autosampler Location: 21  
 Date Collected: 9/5/2024 7:38:01 AM  
 Data Type: Reprocessed on 9/5/2024 8:32:16 AM  
 Initial Sample Vol:  
 Sample Prep Vol:

Nebulizer Parameters: 24-S2465 POST SPK

Analyte	Back Pressure	Flow
All	344.0 kPa	0.65 L/min

-----  
Replicate Data: 24-S2465 POST SPK

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	4388487.5	4387320.8	0.932 mg/L	931.8 ug/L	7:38:53 AM
2	Be 313.107	4355923.2	4354756.6	0.925 mg/L	924.8 ug/L	7:39:10 AM
3	Be 313.107	4400649.9	4399483.3	0.934 mg/L	934.3 ug/L	7:39:27 AM

-----  
Mean Data: 24-S2465 POST SPK

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	4380520.2	0.930 mg/L	0.0049	930.3 ug/L	4.91	0.53%

Sequence No.: 14	Autosampler Location: 22
Sample ID: 24-S2467	Date Collected: 9/5/2024 7:42:20 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:16 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
Nebulizer Parameters: 24-S2467

Analyte	Back Pressure	Flow
All	345.0 kPa	0.65 L/min

-----  
Replicate Data: 24-S2467

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1007.4	-159.3	-0.000 mg/L	-0.034 ug/L	7:43:13 AM
2	Be 313.107	935.1	-231.6	-0.000 mg/L	-0.049 ug/L	7:43:27 AM
3	Be 313.107	919.0	-247.7	-0.000 mg/L	-0.053 ug/L	7:43:41 AM

-----  
Mean Data: 24-S2467

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	-212.8	-0.000 mg/L	0.0000	-0.045 ug/L	0.0100	22.12%

Sequence No.: 15	Autosampler Location: 3
Sample ID: CCV	Date Collected: 9/5/2024 7:46:29 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:16 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
Nebulizer Parameters: CCV

Analyte	Back Pressure	Flow
All	344.0 kPa	0.65 L/min

-----  
Replicate Data: CCV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	11758567.2	11757400.6	2.497 mg/L	2497 ug/L	7:47:22 AM
2	Be 313.107	11744036.7	11742870.0	2.494 mg/L	2494 ug/L	7:47:39 AM
3	Be 313.107	11478341.2	11477174.5	2.437 mg/L	2437 ug/L	7:47:55 AM

-----  
Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	11659148.4	2.476 mg/L	0.0335	2476 ug/L	33.50	1.35%

QC value within limits for Be 313.107 Recovery = 99.04%  
All analyte(s) passed QC.

Sequence No.: 16 Autosampler Location: 4  
Sample ID: CCB Date Collected: 9/5/2024 7:50:46 AM  
Analyst: SG Data Type: Reprocessed on 9/5/2024 8:32:16 AM  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:  
Wash Time:

Nebulizer Parameters: CCB  
Analyte Back Pressure Flow  
All 345.0 kPa 0.65 L/min

Replicate Data: CCB

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1492.9	326.2	0.000 mg/L	0.069 ug/L	7:51:36 AM
2	Be 313.107	1354.2	187.5	0.000 mg/L	0.040 ug/L	7:51:49 AM
3	Be 313.107	1288.9	122.2	0.000 mg/L	0.026 ug/L	7:52:01 AM

Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	212.0	0.000 mg/L	0.0000	0.045 ug/L	0.0221	49.14%

QC value within limits for Be 313.107 Recovery = Not calculated  
All analyte(s) passed QC.

Sequence No.: 17 Autosampler Location: 23  
Sample ID: 24-S2470 Date Collected: 9/5/2024 7:54:50 AM  
Analyst: SG Data Type: Reprocessed on 9/5/2024 8:32:16 AM  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:  
Wash Time:

Nebulizer Parameters: 24-S2470  
Analyte Back Pressure Flow  
All 345.0 kPa 0.65 L/min

Replicate Data: 24-S2470

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1202.0	35.3	0.000 mg/L	0.008 ug/L	7:55:41 AM
2	Be 313.107	1187.6	21.0	0.000 mg/L	0.004 ug/L	7:55:53 AM
3	Be 313.107	1259.9	93.3	0.000 mg/L	0.020 ug/L	7:56:05 AM

Mean Data: 24-S2470

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	49.9	0.000 mg/L	0.0000	0.011 ug/L	0.0081	76.80%

Sequence No.: 18 Autosampler Location: 3  
Sample ID: CCV Date Collected: 9/5/2024 7:58:54 AM  
Analyst: SG Data Type: Reprocessed on 9/5/2024 8:32:16 AM  
Logged In Analyst (Original) : Administrator  
Initial Sample Wt: Initial Sample Vol:  
Dilution: Sample Prep Vol:  
Wash Time:

Nebulizer Parameters: CCV



Analyte Back Pressure Flow  
 All 345.0 kPa 0.65 L/min

-----  
 Replicate Data: CCV

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	11734682.6	11733515.9	2.492 mg/L	2492 ug/L	7:59:46 AM
2	Be 313.107	11793228.8	11792062.1	2.504 mg/L	2504 ug/L	8:00:03 AM
3	Be 313.107	11644707.9	11643541.2	2.473 mg/L	2473 ug/L	8:00:19 AM

-----  
 Mean Data: CCV

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	11723039.8	2.490 mg/L	0.0159	2490 ug/L	15.89	0.64%

QC value within limits for Be 313.107 Recovery = 99.59%  
 All analyte(s) passed QC.

=====

Sequence No.: 19	Autosampler Location: 4
Sample ID: CCB	Date Collected: 9/5/2024 8:03:10 AM
Analyst: SG	Data Type: Reprocessed on 9/5/2024 8:32:16 AM
Logged In Analyst (Original) : Administrator	
Initial Sample Wt:	Initial Sample Vol:
Dilution:	Sample Prep Vol:
Wash Time:	

-----  
 Nebulizer Parameters: CCB

Analyte Back Pressure Flow  
 All 345.0 kPa 0.65 L/min

-----  
 Replicate Data: CCB

Repl#	Analyte	Net Intensity	Corrected Intensity	Calib. Conc. Units	Sample Conc. Units	Analysis Time
1	Be 313.107	1266.5	99.9	0.000 mg/L	0.021 ug/L	8:04:01 AM
2	Be 313.107	1341.3	174.7	0.000 mg/L	0.037 ug/L	8:04:13 AM
3	Be 313.107	1390.1	223.5	0.000 mg/L	0.047 ug/L	8:04:25 AM

-----  
 Mean Data: CCB

Analyte	Mean Corrected Intensity	Calib. Conc. Units	Std.Dev.	Sample Conc. Units	Std.Dev.	RSD
Be 313.107	166.0	0.000 mg/L	0.0000	0.035 ug/L	0.0132	37.50%

QC value within limits for Be 313.107 Recovery = Not calculated  
 All analyte(s) passed QC.